



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY090225\  
Data File : VY023251.D  
Acq On : 02 Sep 2025 08:28  
Operator : SY/MD  
Sample : BFB  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y081225S.M  
Title : SW846 8260  
Last Update : Wed Aug 13 02:10:11 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 18.1         | 32344      | PASS                |
| 75             | 95              | 30              | 60              | 51.4         | 91987      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 178987     | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 11885      | PASS                |
| 173            | 174             | 0.00            | 2               | 1.0          | 1654       | PASS                |
| 174            | 95              | 50              | 100             | 91.5         | 163840     | PASS                |
| 175            | 174             | 5               | 9               | 7.4          | 12173      | PASS                |
| 176            | 174             | 95              | 101             | 95.5         | 156408     | PASS                |
| 177            | 176             | 5               | 9               | 6.5          | 10222      | PASS                |