

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q2994
 Lab File ID: VY023252.D Date Analyzed: 09/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	319834	13.346				
UPPER LIMIT	639668	13.846				
LOWER LIMIT	159917	12.846				
EPA SAMPLE NO.						
PAULDING-WC1	306386	13.35				
VY0902SBL01	430142	13.35				
VY0902SBS01	331529	13.35				
VY0902SBSD01	320346	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.