284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GFEL01

Lab Code: ACE SDG No.: Q3000

Instrument ID: Calibration Date/Time:

Lab File ID: Init. Calib. Date(s):

Heated Purge: (Y/N) Init. Calib. Time(s):

GC Column: ID: (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
				20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.