

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Earth Engineering Inc. Date Collected: 08/29/25 Project: Meetinghouse Date Received: 09/02/25 Client Sample ID: S-4 SDG No.: Q3001 Q3001-04 Lab Sample ID: Matrix: Solid % Solid: 90.1 Analytical Method: **NJEPH** Sample Wt/Vol: 30.09 Final Vol: 2000 Units: uL g Soil Aliquot Vol: иL Test: EPH NF Prep Method:

 Prep Date :
 Date Analyzed :
 Prep Batch ID

 09/03/25 08:39
 09/03/25 20:00
 PB169516

 Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C4	40 Aliphatic C28-C40	4.32		1	1.31	2.21	mg/kg FE055648.1)
Aliphatic C9-C28	8 Aliphatic C9-C28	1.86	J	1	1.01	4.44	mg/kg FE055648.1)
Total AliphaticEl	PH Total AliphaticEPH	6.18	J		2.32	6.65	mg/kg	
Total EPH	Total EPH	6.18	J		2.32	6.65	mg/kg	

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Earth Engineering Inc. Date Collected: 08/29/25 Project: Meetinghouse Date Received: 09/02/25 Client Sample ID: S-4 SDG No.: Q3001 Q3001-04 Lab Sample ID: Matrix: Solid % Solid: 90.1 Analytical Method: **NJEPH** Sample Wt/Vol: 30.09 Final Vol: 2000 Units: uL g Soil Aliquot Vol: иL Test: EPH NF Prep Method:

 Prep Date :
 Date Analyzed :
 Prep Batch ID

 09/03/25 08:39
 09/03/25 20:00
 PB169516

 Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C4	40 Aliphatic C28-C40	4.32		1	1.31	2.21	mg/kg FE055648.1)
Aliphatic C9-C28	8 Aliphatic C9-C28	1.86	J	1	1.01	4.44	mg/kg FE055648.1)
Total AliphaticEl	PH Total AliphaticEPH	6.18	J		2.32	6.65	mg/kg	
Total EPH	Total EPH	6.18	J		2.32	6.65	mg/kg	

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



Final Vol:

2000

uL



Report of Analysis

Client: Earth Engineering Inc. Date Collected: 08/29/25

Project: Meetinghouse Date Received: 09/02/25

Client Sample ID: S-4 SDG No.: Q3001

Lab Sample ID: Q3001-04 Matrix: Solid

Analytical Method: NJEPH % Solid: 90.1 Sample Wt/Vol:

g

30.09

Units:

Soil Aliquot Vol: uL Test: EPH_NF

Prep Method:

File ID: Dilution: Prep Date: Date Analyzed: Prep Batch ID 1 FE055648.D 09/03/25 09/03/25 PB169516

CAS Number	Parameter		Conc. (Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
Aliphatic C9-C	C28	Aliphatic C9-C28	1.86	J	1.01	4.44	mg/kg
Aliphatic C28-	-C40	Aliphatic C28-C40	4.32		1.31	2.21	mg/kg
SURROGATES	8						
3383-33-2		1-chlorooctadecane (SURR)	35.5		40 - 140	71%	SPK: 50
84-15-1		ortho-Terphenyl (SURR)	32.7		40 - 140	65%	SPK: 50



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID: Q3001-04 Acq On: 03 Sep 2025 20:00

Client Sample ID: S-4 Operator: YP\AJ

Data file: FE055648.D Misc:

Instrument: FID_E ALS Vial: 28
Dilution Factor: 1 Sample Multiplier: 1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.314	6.948	606422	4.723	300	ug/ml
Aliphatic C12-C16	6.949	10.401	716801	5.306	200	ug/ml
Aliphatic C16-C21	10.402	13.781	951825	6.6	300	ug/ml
Aliphatic C21-C28	13.782	17.452	1230150	8.588	400	ug/ml
Aliphatic C28-C40	17.453	22.470	8065651	58.566	600	ug/ml
Aliphatic EPH	3.314	22.470	11570849	83.783		ug/ml
ortho-Terphenyl (SURR)	12.075	12.075	5372381	32.7		ug/ml
1-chlorooctadecane (SURR)	13.512	13.512	4433805	35.46		ug/ml
Aliphatic C9-C28	3.314	17.452	3505198	25.217	1200	ug/ml