

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3007

Project: Monthly 2025

Run No.: LB137082

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137082BS							
TPH	mg/L	20.0	17.0		85	1	78-114	09/05/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3007

Project: Monthly 2025

Run No.: LB137082

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137082BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	09/05/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3007

Project: Monthly 2025

Run No.: LB137092

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137092BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	09/05/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3007

Project: Monthly 2025

Run No.: LB137104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137104BS							
TSS	mg/L	550	531		96	1	90-110	09/08/2025