

CASE NARRATIVE

Sciacca General Contractors, LLC
Project Name: 4 Hampton Place, Nutley
Project # N/A
Order ID # Q3011
Test Name: EPH_F2,TPH GC

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 09/03/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:

. This data package contains results for EPH_F2(NJEPH), TPH GC(8015D).

C. Analytical Techniques:

EPH_F2 : The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224.The analysis of EPH_F2s was based on method NJEPH and extraction was done based on method 3541.

TPH GC : The analysis were performed on instrument FID_G. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 13302.The analysis of TPH GC was based on method 8015D and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds except following

EPH_F2 : The MS {Q3011-05MS} with File ID: FE055666.D recoveries met the requirements for all compounds except for Aliphatic [n-Decane (C10)- 187%] due to matrix interference.

TPH GC : The MS {Q2998-01MS} with File ID: FG016549.D recoveries met the requirements for all compounds except for Petroleum Hydrocarbons[-149%] due to matrix interference.

The MSD recoveries met the requirements for all compounds except following

EPH_F2 : The MSD {Q3011-05MSD} with File ID: FE055667.D recoveries met the requirements for all compounds except for Aliphatic [n-Decane (C10)- 197%] due to matrix interference.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

TPH GC : The MSD {Q2998-01MSD} with File ID: FG016550.D recoveries met the requirements for all compounds except for Petroleum Hydrocarbons[-99%] due to matrix interference.

The RPD were met for all analysis except following
TPH GC : The RPD for {Q2998-01MSD} with File ID: FG016550.D met criteria except for Petroleum Hydrocarbons[40.6%] due to difference in MS and MSD concentrations.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____