

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3012

Lab File ID: VY023277.D

BFB Injection Date: 09/04/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:26

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18
75	30.0 - 60.0% of mass 95	50.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.3 (1.4) 1
174	50.0 - 100.0% of mass 95	91.8
175	5.0 - 9.0% of mass 174	6.8 (7.4) 1
176	95.0 - 101.0% of mass 174	87.8 (95.6) 1
177	5.0 - 9.0% of mass 176	5.4 (6.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023278.D	09/04/2025	08:56
VY0904SBL01	VY0904SBL01	VY023279.D	09/04/2025	09:31
VY0904SBS01	VY0904SBS01	VY023280.D	09/04/2025	10:03
VY0904SBSD01	VY0904SBSD01	VY023281.D	09/04/2025	10:26
VOC	Q3012-02	VY023284.D	09/04/2025	13:09