



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ATC Group Services LLC

Project Name: Hazmat survey-Bellevue hospital CPEP

Project # N/A

Order ID # Q3014

Test Name: PCB Group1

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 09/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: PCB Group1. This data package contains results for PCB Group1.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCB Group1s was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PO113755.D met the requirements except for Aroclor-1260(Peak-05) is failing in 1st column, however it is passed in 2nd column therefore no corrective action was taken.

The Continuous Calibration File ID PO113770.D met the requirements except for Aroclor-1260(Peak-04) is failing in 1st column, however it is passed in 2nd column therefore no corrective action was taken.

E. Additional Comments:

The temperature of the samples at the time of receipt was 21.2°C.

Less volume was taken at the time of extraction due to caulk matrix of the samples

No MSMSD performed as samples are caulk matrix.

The soil samples results are based on a dry weight basis.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____