

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3014
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113755.D **Time Analyzed:** 10:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	480.410	500.000	-3.9
Aroclor-1016-2	4.759	4.661	4.861	498.380	500.000	-0.3
Aroclor-1016-3	4.817	4.718	4.918	481.690	500.000	-3.7
Aroclor-1016-4	4.935	4.837	5.037	478.600	500.000	-4.3
Aroclor-1016-5	5.192	5.093	5.293	533.280	500.000	6.7
Aroclor-1260-1	6.226	6.129	6.329	532.660	500.000	6.5
Aroclor-1260-2	6.416	6.319	6.519	508.360	500.000	1.7
Aroclor-1260-3	6.781	6.683	6.883	526.780	500.000	5.4
Aroclor-1260-4	7.039	6.941	7.141	592.200	500.000	18.4
Aroclor-1260-5	7.283	7.185	7.385	613.260	500.000	22.7
Decachlorobiphenyl	8.668	8.570	8.770	54.650	50.000	9.3
Tetrachloro-m-xylene	3.658	3.560	3.760	50.840	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3014
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113755.D **Time Analyzed:** 10:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	467.370	500.000	-6.5
Aroclor-1016-2	4.739	4.641	4.841	478.470	500.000	-4.3
Aroclor-1016-3	4.913	4.816	5.016	476.440	500.000	-4.7
Aroclor-1016-4	4.956	4.858	5.058	480.270	500.000	-3.9
Aroclor-1016-5	5.167	5.070	5.270	492.560	500.000	-1.5
Aroclor-1260-1	6.193	6.096	6.296	477.990	500.000	-4.4
Aroclor-1260-2	6.381	6.284	6.484	469.900	500.000	-6.0
Aroclor-1260-3	6.532	6.435	6.635	456.560	500.000	-8.7
Aroclor-1260-4	7.001	6.903	7.103	478.160	500.000	-4.4
Aroclor-1260-5	7.244	7.146	7.346	471.500	500.000	-5.7
Decachlorobiphenyl	8.612	8.514	8.714	53.250	50.000	6.5
Tetrachloro-m-xylene	3.649	3.551	3.751	49.190	50.000	-1.6