



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3014	Analytical Method:	8082A
Client:	ATC Group Services LLC	Datafile :	PO113761.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169705BS (Column 1)	AR1016	166.6	146	ug/kg	88				71	120	
	AR1260	166.6	158	ug/kg	95				65	130	
	Total PCBs	0	278	ug/kg					0	0	
PB169705BS (Column 2)	AR1016	166.6	141	ug/kg	85				71	120	
	AR1260	166.6	137	ug/kg	82				65	130	
	Total PCBs	0	304	ug/kg					0	0	