



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Alliance Technical Group, LLC - Newark**

**Project Name: NJ Waste Water PT**

**Project # N/A**

**Order ID # Q3015**

**Test Name: SVOCMS Group3**

### **A. Number of Samples and Date of Receipt:**

1 Water sample was received on 09/04/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group3. This data package contains results for SVOCMS Group3.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOCMS Group3 was based on method 8270-Modified and extraction was done based on method 3510.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for WP0925-PT-BN-WP [2-Methylnaphthalene-d10 - 0%, Fluoranthene-d10 - 0%], WP0925-PT-BN-WPDL [2-Methylnaphthalene-d10 - 0%, Fluoranthene-d10 - 0%], These samples were extracted for full scan analysis and above mention surrogates were not part of full scan extraction therefore no corrective action is required.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration Met the Requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Sample WP0925-PT-BN-WP was diluted due to high concentration.

### **E. Additional Comments:**



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_