



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Alliance Technical Group, LLC - Newark

Project Name: NJ Waste Water PT

Project # N/A

Order ID # Q3015

Test Name: SVOCMS Group4

A. Number of Samples and Date of Receipt:

1 Water sample was received on 09/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group4. This data package contains results for SVOCMS Group4.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOCMS Group4 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for WP0925-PT-ACIDS-WP [2,4,6-Tribromophenol - 213%, 2-Fluorophenol - 124%, Phenol-d6 - 131%, Terphenyl-d14 - 37016%], WP0925-PT-ACIDS-WPDL [2-Fluorophenol - 111% and Phenol-d6 - 113%], These samples were extracted for full scan analysis and above mention surrogates were not part of full scan extraction therefore no corrective action is required.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for {PB169712BSD} with File ID: BN037888.D met criteria except for Pentachlorophenol[36%] , due to difference in results of BS and BSD.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The %RSD is greater than 20% in the Method 8270-SIM-BN092325.M for 4,6-Dinitro-2-methylphenol and this compound is passing on Quadratic regression.removed from 5.0 ppm.

The Continuous Calibration criateria met the Requirements.
The Tuning criteria met requirements.

Sample WP0925-PT-ACIDS-WP was diluted due to high concentration.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____