

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: ALLI03
ProjectID: NJ Waste Water PT
Lab Code: ACE SDG No.: Q3015
DataFile: FG016624.D Analyst Name: YP\AJ Analyst Date: 09-18-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	57688481	115377	124148	7.065