

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:		
Project:	NJ Waste Water PT		Date Received:		
Client Sample ID:	BSF0917W1		SDG No.:	Q3015	
Lab Sample ID:	BSF0917W1		Matrix:	Water	
Analytical Method:	8015D GRO		% Solid:	0	Decanted:
Sample Wt/Vol:	5	Units: mL	Final Vol:	5	mL
Soil Aliquot Vol:		uL	Test:	Gasoline Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :	PH :				
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
FB032200.D	1	09/17/25 12:45	FB091725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
GRO	GRO	194		6.00	45.0	ug/L
SURROGATES						
98-08-8	Alpha,Alpha,Alpha-Trifluoroto	20.4		50 - 150	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit