

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 17:34

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.47	7.47	7.37	7.57	0.00
MCP	7.65	7.65	7.55	7.75	0.00
2,4-DCAA	7.29	7.29	7.19	7.39	0.01
Dalapon	2.67	2.67	2.57	2.77	0.00
MCPA	7.80	7.80	7.70	7.90	0.00
DICHLORPROP	8.18	8.18	8.08	8.28	0.00
2,4-D	8.42	8.41	8.31	8.51	-0.01
2,4,5-TP(Silvex)	9.29	9.29	9.19	9.39	0.00
2,4,5-T	9.58	9.58	9.48	9.68	0.00
2,4-DB	10.16	10.16	10.06	10.26	0.00
Dinoseb	11.36	11.35	11.25	11.45	-0.01
Pentachlorophenol	8.71	8.71	8.61	8.81	0.00
4-Nitrophenol	7.10	7.10	7.00	7.20	0.00
PICLORAM	11.19	11.18	11.08	11.28	-0.01
DCPA	11.65	11.65	11.55	11.75	0.00
3,5-DICHLOROBENZ	6.45	6.45	6.35	6.55	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 17:34

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.94	7.94	7.84	8.04	0.00
MCPD	8.04	8.04	7.94	8.14	0.00
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
Dalapon	2.69	2.69	2.59	2.79	0.00
MCPA	8.28	8.28	8.18	8.38	0.00
DICHLORPROP	8.66	8.66	8.56	8.76	0.00
2,4-D	9.00	9.00	8.90	9.10	0.00
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00
2,4,5-T	10.32	10.32	10.22	10.42	0.00
2,4-DB	10.89	10.89	10.79	10.99	0.00
Dinoseb	11.27	11.27	11.17	11.37	0.00
Pentachlorophenol	9.51	9.51	9.41	9.61	0.00
4-Nitrophenol	7.30	7.29	7.19	7.39	-0.01
PICLORAM	12.39	12.39	12.29	12.49	0.00
DCPA	12.31	12.31	12.21	12.41	0.00
3,5-DICHLOROBENZ	6.70	6.70	6.60	6.80	0.00