

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031675.D **Time Analyzed:** 14:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.583	9.483	9.683	709.640	712.500	-0.4
2,4,5-TP(Silvex)	9.286	9.185	9.385	700.910	712.500	-1.6
2,4-D	8.415	8.314	8.514	673.870	705.000	-4.4
2,4-DB	10.159	10.058	10.258	699.830	712.500	-1.8
2,4-DCAA	7.286	7.185	7.385	743.530	750.000	-0.9
3,5-DICHLOROBENZOIC ACID	6.453	6.353	6.553	683.140	697.500	-2.1
4-Nitrophenol	7.097	6.995	7.195	716.620	682.500	5.0
Dalapon	2.670	2.569	2.769	665.380	682.500	-2.5
DCPA	11.654	11.554	11.754	710.430	720.000	-1.3
DICAMBA	7.469	7.369	7.569	703.960	705.000	-0.1
DICHLORPROP	8.178	8.077	8.277	683.240	705.000	-3.1
Dinoseb	11.355	11.254	11.454	729.410	705.000	3.5
MCPA	7.796	7.696	7.896	69.930	69.750	0.3
MCPP	7.647	7.547	7.747	73.010	70.500	3.6
Pentachlorophenol	8.708	8.608	8.808	751.190	712.500	5.4
PICLORAM	11.185	11.084	11.284	696.550	712.500	-2.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031675.D **Time Analyzed:** 14:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.322	10.221	10.421	697.640	712.500	-2.1
2,4,5-TP(Silvex)	9.894	9.793	9.993	703.340	712.500	-1.3
2,4-D	9.001	8.900	9.100	683.870	705.000	-3.0
2,4-DB	10.891	10.790	10.990	656.790	712.500	-7.8
2,4-DCAA	7.743	7.643	7.843	724.200	750.000	-3.4
3,5-DICHLOROBENZOIC ACID	6.696	6.595	6.795	695.520	697.500	-0.3
4-Nitrophenol	7.294	7.191	7.391	669.640	682.500	-1.9
Dalapon	2.692	2.591	2.791	675.600	682.500	-1.0
DCPA	12.306	12.207	12.407	717.150	720.000	-0.4
DICAMBA	7.939	7.839	8.039	693.840	705.000	-1.6
DICHLORPROP	8.659	8.559	8.759	706.990	705.000	0.3
Dinoseb	11.266	11.166	11.366	747.350	705.000	6.0
MCPA	8.284	8.184	8.384	71.980	69.750	3.2
MCPP	8.036	7.936	8.136	76.050	70.500	7.9
Pentachlorophenol	9.509	9.409	9.609	727.250	712.500	2.1
PICLORAM	12.387	12.285	12.485	719.220	712.500	0.9