

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031683.D **Time Analyzed:** 17:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.584	9.483	9.683	728.820	712.500	2.3
2,4,5-TP(Silvex)	9.286	9.185	9.385	713.210	712.500	0.1
2,4-D	8.416	8.314	8.514	685.790	705.000	-2.7
2,4-DB	10.159	10.058	10.258	722.740	712.500	1.4
2,4-DCAA	7.285	7.185	7.385	756.390	750.000	0.9
3,5-DICHLOROBENZOIC ACID	6.454	6.353	6.553	691.760	697.500	-0.8
4-Nitrophenol	7.097	6.995	7.195	752.370	682.500	10.2
Dalapon	2.670	2.569	2.769	672.870	682.500	-1.4
DCPA	11.654	11.554	11.754	726.600	720.000	0.9
DICAMBA	7.470	7.369	7.569	711.750	705.000	1.0
DICHLORPROP	8.178	8.077	8.277	693.630	705.000	-1.6
Dinoseb	11.356	11.254	11.454	758.780	705.000	7.6
MCPA	7.797	7.696	7.896	71.360	69.750	2.3
MCPP	7.647	7.547	7.747	73.760	70.500	4.6
Pentachlorophenol	8.709	8.608	8.808	758.790	712.500	6.5
PICLORAM	11.186	11.084	11.284	720.280	712.500	1.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031683.D **Time Analyzed:** 17:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.323	10.221	10.421	703.170	712.500	-1.3
2,4,5-TP(Silvex)	9.894	9.793	9.993	709.360	712.500	-0.4
2,4-D	9.002	8.900	9.100	686.600	705.000	-2.6
2,4-DB	10.891	10.790	10.990	673.830	712.500	-5.4
2,4-DCAA	7.744	7.643	7.843	727.340	750.000	-3.0
3,5-DICHLOROBENZOIC ACID	6.696	6.595	6.795	706.160	697.500	1.2
4-Nitrophenol	7.295	7.191	7.391	685.130	682.500	0.4
Dalapon	2.691	2.591	2.791	682.240	682.500	0.0
DCPA	12.307	12.207	12.407	727.520	720.000	1.0
DICAMBA	7.940	7.839	8.039	700.110	705.000	-0.7
DICHLORPROP	8.659	8.559	8.759	712.230	705.000	1.0
Dinoseb	11.267	11.166	11.366	779.110	705.000	10.5
MCPA	8.284	8.184	8.384	72.470	69.750	3.9
MCPP	8.036	7.936	8.136	76.270	70.500	8.2
Pentachlorophenol	9.509	9.409	9.609	735.530	712.500	3.2
PICLORAM	12.387	12.285	12.485	723.470	712.500	1.5