

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/08/2025

09/08/2025

Continuing Calib Time: 00:11

Initial Calibration Time(s): 20:32

21:46

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1232-1 (1)	4.03	4.03	3.93	4.13	0.00
Aroclor-1232-2 (2)	4.52	4.52	4.42	4.62	0.00
Aroclor-1232-3 (3)	4.76	4.76	4.66	4.86	0.00
Aroclor-1232-4 (4)	4.94	4.94	4.84	5.04	0.00
Aroclor-1232-5 (5)	4.98	4.98	4.88	5.08	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.67	8.67	8.57	8.77	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/08/2025

09/08/2025

Continuing Calib Time: 00:11

Initial Calibration Time(s): 20:32

21:46

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1232-1 (1)	4.02	4.02	3.92	4.12	0.00
Aroclor-1232-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1232-3 (3)	4.91	4.91	4.81	5.01	0.00
Aroclor-1232-4 (4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1232-5 (5)	5.17	5.17	5.07	5.27	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.62	8.62	8.52	8.72	0.01