

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/08/2025

09/08/2025

Continuing Calib Time: 02:57

Initial Calibration Time(s): 20:32

21:46

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1232-1 (1)   | 4.03       | 4.03      | 3.93      | 4.13 | 0.00       |
| Aroclor-1232-2 (2)   | 4.52       | 4.52      | 4.42      | 4.62 | 0.00       |
| Aroclor-1232-3 (3)   | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1232-4 (4)   | 4.94       | 4.94      | 4.84      | 5.04 | 0.01       |
| Aroclor-1232-5 (5)   | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.67       | 8.67      | 8.57      | 8.77 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/09/2025

Initial Calibration Date(s): 09/08/2025

09/08/2025

Continuing Calib Time: 02:57

Initial Calibration Time(s): 20:32

21:46

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1232-1 (1)   | 4.02       | 4.02      | 3.92      | 4.12 | 0.00       |
| Aroclor-1232-2 (2)   | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1232-3 (3)   | 4.91       | 4.91      | 4.81      | 5.01 | 0.00       |
| Aroclor-1232-4 (4)   | 5.00       | 5.00      | 4.90      | 5.10 | 0.00       |
| Aroclor-1232-5 (5)   | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Tetrachloro-m-xylene | 3.65       | 3.65      | 3.55      | 3.75 | 0.00       |
| Decachlorobiphenyl   | 8.61       | 8.62      | 8.52      | 8.72 | 0.01       |