

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q3015  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 09/08/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113568.D **Time Analyzed:** 11:29

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.742 | 4.648     | 4.848 | 533.910            | 500.000           | 6.8  |
| Aroclor-1016-2       | 4.760 | 4.666     | 4.866 | 542.900            | 500.000           | 8.6  |
| Aroclor-1016-3       | 4.817 | 4.724     | 4.924 | 540.010            | 500.000           | 8.0  |
| Aroclor-1016-4       | 4.935 | 4.843     | 5.043 | 509.940            | 500.000           | 2.0  |
| Aroclor-1016-5       | 5.192 | 5.099     | 5.299 | 542.130            | 500.000           | 8.4  |
| Aroclor-1260-1       | 6.227 | 6.136     | 6.336 | 534.580            | 500.000           | 6.9  |
| Aroclor-1260-2       | 6.418 | 6.326     | 6.526 | 536.840            | 500.000           | 7.4  |
| Aroclor-1260-3       | 6.782 | 6.691     | 6.891 | 551.850            | 500.000           | 10.4 |
| Aroclor-1260-4       | 7.041 | 6.951     | 7.151 | 543.830            | 500.000           | 8.8  |
| Aroclor-1260-5       | 7.284 | 7.194     | 7.394 | 540.000            | 500.000           | 8.0  |
| Decachlorobiphenyl   | 8.669 | 8.582     | 8.782 | 50.110             | 50.000            | 0.2  |
| Tetrachloro-m-xylene | 3.658 | 3.563     | 3.763 | 56.250             | 50.000            | 12.5 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q3015  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 09/08/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113568.D **Time Analyzed:** 11:29

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.722 | 4.629     | 4.829 | 497.870            | 500.000           | -0.4  |
| Aroclor-1016-2       | 4.739 | 4.647     | 4.847 | 513.360            | 500.000           | 2.7   |
| Aroclor-1016-3       | 4.914 | 4.822     | 5.022 | 464.210            | 500.000           | -7.2  |
| Aroclor-1016-4       | 4.957 | 4.864     | 5.064 | 461.890            | 500.000           | -7.6  |
| Aroclor-1016-5       | 5.168 | 5.076     | 5.276 | 525.760            | 500.000           | 5.2   |
| Aroclor-1260-1       | 6.195 | 6.104     | 6.304 | 478.190            | 500.000           | -4.4  |
| Aroclor-1260-2       | 6.383 | 6.292     | 6.492 | 461.890            | 500.000           | -7.6  |
| Aroclor-1260-3       | 6.534 | 6.443     | 6.643 | 442.080            | 500.000           | -11.6 |
| Aroclor-1260-4       | 7.002 | 6.912     | 7.112 | 454.990            | 500.000           | -9.0  |
| Aroclor-1260-5       | 7.245 | 7.155     | 7.355 | 445.410            | 500.000           | -10.9 |
| Decachlorobiphenyl   | 8.614 | 8.528     | 8.728 | 52.820             | 50.000            | 5.6   |
| Tetrachloro-m-xylene | 3.649 | 3.555     | 3.755 | 52.860             | 50.000            | 5.7   |