

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113580.D **Time Analyzed:** 17:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.648	4.848	546.720	500.000	9.3
Aroclor-1016-2	4.761	4.666	4.866	556.800	500.000	11.4
Aroclor-1016-3	4.818	4.724	4.924	558.620	500.000	11.7
Aroclor-1016-4	4.936	4.843	5.043	530.220	500.000	6.0
Aroclor-1016-5	5.194	5.099	5.299	562.270	500.000	12.5
Aroclor-1260-1	6.229	6.136	6.336	549.030	500.000	9.8
Aroclor-1260-2	6.419	6.326	6.526	546.010	500.000	9.2
Aroclor-1260-3	6.784	6.691	6.891	546.610	500.000	9.3
Aroclor-1260-4	7.043	6.951	7.151	541.750	500.000	8.4
Aroclor-1260-5	7.287	7.194	7.394	536.430	500.000	7.3
Decachlorobiphenyl	8.672	8.582	8.782	50.160	50.000	0.3
Tetrachloro-m-xylene	3.659	3.563	3.763	56.040	50.000	12.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113580.D **Time Analyzed:** 17:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.629	4.829	497.300	500.000	-0.5
Aroclor-1016-2	4.741	4.647	4.847	510.520	500.000	2.1
Aroclor-1016-3	4.915	4.822	5.022	476.940	500.000	-4.6
Aroclor-1016-4	4.958	4.864	5.064	472.810	500.000	-5.4
Aroclor-1016-5	5.170	5.076	5.276	530.680	500.000	6.1
Aroclor-1260-1	6.196	6.104	6.304	472.050	500.000	-5.6
Aroclor-1260-2	6.384	6.292	6.492	452.870	500.000	-9.4
Aroclor-1260-3	6.535	6.443	6.643	430.880	500.000	-13.8
Aroclor-1260-4	7.004	6.912	7.112	437.790	500.000	-12.4
Aroclor-1260-5	7.247	7.155	7.355	424.710	500.000	-15.1
Decachlorobiphenyl	8.616	8.528	8.728	52.970	50.000	5.9
Tetrachloro-m-xylene	3.651	3.555	3.755	51.760	50.000	3.5