

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113593.D **Time Analyzed:** 22:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.742	4.648	4.848	531.910	500.000	6.4
Aroclor-1016-2	4.760	4.666	4.866	548.630	500.000	9.7
Aroclor-1016-3	4.818	4.724	4.924	540.400	500.000	8.1
Aroclor-1016-4	4.936	4.843	5.043	516.110	500.000	3.2
Aroclor-1016-5	5.193	5.099	5.299	559.230	500.000	11.8
Aroclor-1260-1	6.229	6.136	6.336	534.630	500.000	6.9
Aroclor-1260-2	6.418	6.326	6.526	542.440	500.000	8.5
Aroclor-1260-3	6.783	6.691	6.891	543.170	500.000	8.6
Aroclor-1260-4	7.042	6.951	7.151	543.690	500.000	8.7
Aroclor-1260-5	7.285	7.194	7.394	540.690	500.000	8.1
Decachlorobiphenyl	8.670	8.582	8.782	52.300	50.000	4.6
Tetrachloro-m-xylene	3.659	3.563	3.763	55.100	50.000	10.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113593.D **Time Analyzed:** 22:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.629	4.829	494.430	500.000	-1.1
Aroclor-1016-2	4.740	4.647	4.847	503.470	500.000	0.7
Aroclor-1016-3	4.915	4.822	5.022	511.390	500.000	2.3
Aroclor-1016-4	4.958	4.864	5.064	496.400	500.000	-0.7
Aroclor-1016-5	5.169	5.076	5.276	538.260	500.000	7.7
Aroclor-1260-1	6.195	6.104	6.304	470.440	500.000	-5.9
Aroclor-1260-2	6.384	6.292	6.492	456.610	500.000	-8.7
Aroclor-1260-3	6.534	6.443	6.643	429.290	500.000	-14.1
Aroclor-1260-4	7.003	6.912	7.112	438.450	500.000	-12.3
Aroclor-1260-5	7.246	7.155	7.355	422.410	500.000	-15.5
Decachlorobiphenyl	8.614	8.528	8.728	54.250	50.000	8.5
Tetrachloro-m-xylene	3.650	3.555	3.755	52.390	50.000	4.8