

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/08/2025 09/08/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1232CCC500 **Data File :** PO113594.D **Time Analyzed:** 00:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1232-1	4.031	3.930	4.130	527.120	500.000	5.4
Aroclor-1232-2	4.521	4.421	4.621	486.040	500.000	-2.8
Aroclor-1232-3	4.760	4.660	4.860	545.950	500.000	9.2
Aroclor-1232-4	4.936	4.836	5.036	564.480	500.000	12.9
Aroclor-1232-5	4.981	4.879	5.079	480.240	500.000	-4.0
Decachlorobiphenyl	8.671	8.570	8.770	53.610	50.000	7.2
Tetrachloro-m-xylene	3.659	3.558	3.758	58.460	50.000	16.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/08/2025 09/08/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1232CCC500 **Data File :** PO113594.D **Time Analyzed:** 00:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1232-1	4.018	3.917	4.117	526.070	500.000	5.2
Aroclor-1232-2	4.740	4.640	4.840	532.890	500.000	6.6
Aroclor-1232-3	4.914	4.814	5.014	518.180	500.000	3.6
Aroclor-1232-4	4.999	4.898	5.098	498.590	500.000	-0.3
Aroclor-1232-5	5.168	5.069	5.269	497.370	500.000	-0.5
Decachlorobiphenyl	8.615	8.515	8.715	55.700	50.000	11.4
Tetrachloro-m-xylene	3.650	3.549	3.749	55.260	50.000	10.5