

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113598.D **Time Analyzed:** 02:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.648	4.848	541.170	500.000	8.2
Aroclor-1016-2	4.760	4.666	4.866	570.360	500.000	14.1
Aroclor-1016-3	4.817	4.724	4.924	569.040	500.000	13.8
Aroclor-1016-4	4.936	4.843	5.043	547.240	500.000	9.4
Aroclor-1016-5	5.193	5.099	5.299	607.990	500.000	21.6
Aroclor-1260-1	6.227	6.136	6.336	555.640	500.000	11.1
Aroclor-1260-2	6.417	6.326	6.526	563.480	500.000	12.7
Aroclor-1260-3	6.782	6.691	6.891	561.490	500.000	12.3
Aroclor-1260-4	7.040	6.951	7.151	559.130	500.000	11.8
Aroclor-1260-5	7.284	7.194	7.394	554.270	500.000	10.9
Decachlorobiphenyl	8.669	8.582	8.782	53.130	50.000	6.3
Tetrachloro-m-xylene	3.658	3.563	3.763	57.330	50.000	14.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113598.D **Time Analyzed:** 02:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.629	4.829	509.650	500.000	1.9
Aroclor-1016-2	4.740	4.647	4.847	517.470	500.000	3.5
Aroclor-1016-3	4.914	4.822	5.022	537.490	500.000	7.5
Aroclor-1016-4	4.957	4.864	5.064	514.400	500.000	2.9
Aroclor-1016-5	5.168	5.076	5.276	569.490	500.000	13.9
Aroclor-1260-1	6.195	6.104	6.304	480.900	500.000	-3.8
Aroclor-1260-2	6.383	6.292	6.492	468.980	500.000	-6.2
Aroclor-1260-3	6.534	6.443	6.643	435.360	500.000	-12.9
Aroclor-1260-4	7.002	6.912	7.112	445.620	500.000	-10.9
Aroclor-1260-5	7.245	7.155	7.355	433.310	500.000	-13.3
Decachlorobiphenyl	8.614	8.528	8.728	54.970	50.000	9.9
Tetrachloro-m-xylene	3.649	3.555	3.755	54.290	50.000	8.6