

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/08/2025 09/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1232CCC500 **Data File :** PO113599.D **Time Analyzed:** 02:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1232-1	4.030	3.930	4.130	517.770	500.000	3.6
Aroclor-1232-2	4.520	4.421	4.621	494.720	500.000	-1.1
Aroclor-1232-3	4.760	4.660	4.860	524.340	500.000	4.9
Aroclor-1232-4	4.935	4.836	5.036	528.140	500.000	5.6
Aroclor-1232-5	4.979	4.879	5.079	484.930	500.000	-3.0
Decachlorobiphenyl	8.668	8.570	8.770	52.850	50.000	5.7
Tetrachloro-m-xylene	3.658	3.558	3.758	57.360	50.000	14.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/08/2025 09/08/2025

Client Sample No.: CCAL06 **Date Analyzed:** 09/09/2025
Lab Sample No.: AR1232CCC500 **Data File :** PO113599.D **Time Analyzed:** 02:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1232-1	4.017	3.917	4.117	519.510	500.000	3.9
Aroclor-1232-2	4.739	4.640	4.840	521.790	500.000	4.4
Aroclor-1232-3	4.914	4.814	5.014	539.520	500.000	7.9
Aroclor-1232-4	4.998	4.898	5.098	538.660	500.000	7.7
Aroclor-1232-5	5.168	5.069	5.269	540.520	500.000	8.1
Decachlorobiphenyl	8.614	8.515	8.715	54.390	50.000	8.8
Tetrachloro-m-xylene	3.649	3.549	3.749	52.360	50.000	4.7