



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8082A

Client: Alliance Technical Group, LLC - Newark

Datafile : PO113576.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169600BSD (Column 1)	AR1016	5	4.60	ug/L	92	0			77	107	20
	AR1260	5	4.50	ug/L	90	2			66	113	20
PB169600BSD (Column 2)	AR1016	5	4.20	ug/L	84	0			77	107	20
	AR1260	5	3.80	ug/L	76	0			66	113	20