

Surrogate Summary

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089961.D	PIBLK-PD089961.D	Decachlorobiphen	1	20	19.3	96		57	171
		Tetrachloro-m-xyl	1	20	17.4	87		61	148
		Decachlorobiphen	2	20	18.3	91		57	171
		Tetrachloro-m-xyl	2	20	17.5	88		61	148
I.BLK-PD090194.D	PIBLK-PD090194.D	Decachlorobiphen	1	20	18.1	90		57	171
		Tetrachloro-m-xyl	1	20	19.0	95		61	148
		Decachlorobiphen	2	20	14.9	74		57	171
		Tetrachloro-m-xyl	2	20	20.6	103		61	148
PB169601BL	PB169601BL	Decachlorobiphen	1	20	16.3	81		57	171
		Tetrachloro-m-xyl	1	20	16.8	84		61	148
		Decachlorobiphen	2	20	14.4	72		57	171
		Tetrachloro-m-xyl	2	20	18.4	92		61	148
Q3015-09	WP0925-PT-PEST-WP	Decachlorobiphen	1	20	17.8	89		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	16.1	80		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
Q3015-09DL	WP0925-PT-PEST-WPDL	Decachlorobiphen	1	20	21.4	107		57	171
		Tetrachloro-m-xyl	1	20	24.0	120		61	148
		Decachlorobiphen	2	20	19.5	98		57	171
		Tetrachloro-m-xyl	2	20	25.6	128		61	148
Q3015-09DL2	WP0925-PT-PEST-WPDL2	Decachlorobiphen	1	20	0	0	*	57	171
		Tetrachloro-m-xyl	1	20	0	0	*	61	148
		Decachlorobiphen	2	20	0	0	*	57	171
		Tetrachloro-m-xyl	2	20	0	0	*	61	148
PB169601BS	PB169601BS	Decachlorobiphen	1	20	19.3	96		57	171
		Tetrachloro-m-xyl	1	20	19.3	97		61	148
		Decachlorobiphen	2	20	18.1	90		57	171
		Tetrachloro-m-xyl	2	20	20.8	104		61	148
PB169601BSD	PB169601BSD	Decachlorobiphen	1	20	19.6	98		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	18.1	91		57	171
		Tetrachloro-m-xyl	2	20	20.6	103		61	148
I.BLK-PD090202.D	PIBLK-PD090202.D	Decachlorobiphen	1	20	19.0	95		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	18.2	91		57	171
		Tetrachloro-m-xyl	2	20	20.6	103		61	148