

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8081B

Client: Alliance Technical Group, LLC - Newark

Datafile : PD090201.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169601BSD (Column 1)	alpha-BHC	0.5	0.56	ug/L	113	1			85	130	20
	beta-BHC	0.5	0.55	ug/L	110	1			83	126	20
	delta-BHC	0.5	0.46	ug/L	93	1			69	141	20
	gamma-BHC (Lindane)	0.5	0.56	ug/L	113	1			82	129	20
	Heptachlor	0.5	0.55	ug/L	110	1			79	127	20
	Aldrin	0.5	0.58	ug/L	117	1			79	126	20
	Heptachlor epoxide	0.5	0.58	ug/L	115	1			81	124	20
	Endosulfan I	0.5	0.57	ug/L	113	1			85	122	20
	Dieldrin	0.5	0.57	ug/L	113	2			83	125	20
	4,4'-DDE	0.5	0.57	ug/L	114	0			80	127	20
	Endrin	0.5	0.48	ug/L	97	0			81	128	20
	Endosulfan II	0.5	0.56	ug/L	111	1			82	123	20
	4,4'-DDD	0.5	0.61	ug/L	121	0			77	131	20
	Endosulfan sulfate	0.5	0.51	ug/L	103	1			76	129	20
	4,4'-DDT	0.5	0.47	ug/L	94	1			80	133	20
	Methoxychlor	0.5	0.44	ug/L	87	0			78	108	20
	Endrin ketone	0.5	0.57	ug/L	114	1			80	131	20
	Endrin aldehyde	0.5	0.61	ug/L	122	1			82	127	20
	alpha-Chlordane	0.5	0.57	ug/L	114	0			82	125	20
	gamma-Chlordane	0.5	0.58	ug/L	116	0			82	125	20
PB169601BSD (Column 2)	alpha-BHC	0.5	0.56	ug/L	113	0			85	130	20
	beta-BHC	0.5	0.57	ug/L	115	0			83	126	20
	delta-BHC	0.5	0.44	ug/L	88	0			69	141	20
	gamma-BHC (Lindane)	0.5	0.57	ug/L	114	0			82	129	20
	Heptachlor	0.5	0.55	ug/L	110	0			79	127	20
	Aldrin	0.5	0.58	ug/L	116	1			79	126	20
	Heptachlor epoxide	0.5	0.58	ug/L	116	0			81	124	20
	Endosulfan I	0.5	0.57	ug/L	114	1			85	122	20
	Dieldrin	0.5	0.58	ug/L	116	0			83	125	20
	4,4'-DDE	0.5	0.59	ug/L	117	0			80	127	20
	Endrin	0.5	0.53	ug/L	105	0			81	128	20
	Endosulfan II	0.5	0.57	ug/L	114	0			82	123	20
	4,4'-DDD	0.5	0.61	ug/L	121	1			77	131	20
	Endosulfan sulfate	0.5	0.53	ug/L	105	1			76	129	20
	4,4'-DDT	0.5	0.50	ug/L	100	1			80	133	20
	Methoxychlor	0.5	0.48	ug/L	96	2			78	108	20
	Endrin ketone	0.5	0.57	ug/L	113	2			80	131	20
	Endrin aldehyde	0.5	0.63	ug/L	126	0			82	127	20
	alpha-Chlordane	0.5	0.58	ug/L	116	0			82	125	20
	gamma-Chlordane	0.5	0.59	ug/L	117	0			82	125	20