

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/10/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 15:41

Initial Calibration Time(s): 10:30

11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Chlordane-1 (1)	4.71	4.71	4.61	4.81	0.00
Chlordane-2 (2)	5.24	5.24	5.14	5.34	0.00
Chlordane-3 (3)	5.95	5.94	5.84	6.04	-0.01
Chlordane-4 (4)	6.03	6.03	5.93	6.13	0.00
Chlordane-5 (5)	6.87	6.87	6.77	6.97	0.00
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Continuing Calib Date: 09/10/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 15:41

Initial Calibration Time(s): 10:30

11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Chlordane-1 (1)	3.90	3.90	3.80	4.00	0.00
Chlordane-2 (2)	4.48	4.48	4.38	4.58	0.00
Chlordane-3 (3)	5.12	5.12	5.02	5.22	0.00
Chlordane-4 (4)	5.18	5.18	5.08	5.28	0.00
Chlordane-5 (5)	6.08	6.08	5.98	6.18	0.00
Decachlorobiphenyl	8.07	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00