

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD090222.D **Time Analyzed:** 15:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.713	4.613	4.813	474.820	500.000	-5.0
Chlordane-2	5.242	5.139	5.339	519.770	500.000	4.0
Chlordane-3	5.947	5.844	6.044	520.610	500.000	4.1
Chlordane-4	6.030	5.929	6.129	532.410	500.000	6.5
Chlordane-5	6.871	6.768	6.968	552.610	500.000	10.5
Decachlorobiphenyl	9.071	8.968	9.168	42.200	50.000	-15.6
Tetrachloro-m-xylene	3.550	3.449	3.649	42.140	50.000	-15.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD090222.D **Time Analyzed:** 15:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.900	3.801	4.001	581.700	500.000	16.3
Chlordane-2	4.481	4.382	4.582	552.020	500.000	10.4
Chlordane-3	5.120	5.020	5.220	580.760	500.000	16.2
Chlordane-4	5.184	5.084	5.284	573.250	500.000	14.7
Chlordane-5	6.084	5.984	6.184	590.240	500.000	18.0
Decachlorobiphenyl	8.065	7.964	8.164	45.470	50.000	-9.1
Tetrachloro-m-xylene	2.875	2.776	2.976	58.390	50.000	16.8