

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/10/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD090234.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.709	4.613	4.813	541.810	500.000	8.4
Chlordane-2	5.236	5.139	5.339	515.740	500.000	3.1
Chlordane-3	5.941	5.844	6.044	531.370	500.000	6.3
Chlordane-4	6.027	5.929	6.129	515.080	500.000	3.0
Chlordane-5	6.866	6.768	6.968	545.600	500.000	9.1
Decachlorobiphenyl	9.066	8.968	9.168	41.150	50.000	-17.7
Tetrachloro-m-xylene	3.546	3.449	3.649	42.490	50.000	-15.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q3015
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/10/2025
Lab Sample No.: PCHLORCCC500 **Data File :** PD090234.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.900	3.801	4.001	584.670	500.000	16.9
Chlordane-2	4.481	4.382	4.582	557.170	500.000	11.4
Chlordane-3	5.119	5.020	5.220	582.070	500.000	16.4
Chlordane-4	5.183	5.084	5.284	569.480	500.000	13.9
Chlordane-5	6.083	5.984	6.184	590.220	500.000	18.0
Decachlorobiphenyl	8.064	7.964	8.164	45.570	50.000	-8.9
Tetrachloro-m-xylene	2.876	2.776	2.976	57.490	50.000	15.0