

Surrogate Summary

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089983.D	PIBLK-PD089983.D	Decachlorobiphen	1	20	20.6	103		57	171
		Tetrachloro-m-xyl	1	20	18.7	94		61	148
		Decachlorobiphen	2	20	21.4	107		57	171
		Tetrachloro-m-xyl	2	20	20.2	101		61	148
I.BLK-PD090219.D	PIBLK-PD090219.D	Decachlorobiphen	1	20	19.3	97		57	171
		Tetrachloro-m-xyl	1	20	17.9	89		61	148
		Decachlorobiphen	2	20	20.1	101		57	171
		Tetrachloro-m-xyl	2	20	20.3	101		61	148
Q3015-11	WP0925-PT-CHLR-WP	Decachlorobiphen	1	20	22.8	114		57	171
		Tetrachloro-m-xyl	1	20	19.7	99		61	148
		Decachlorobiphen	2	20	24.1	120		57	171
		Tetrachloro-m-xyl	2	20	27.4	137		61	148
PB169636BL	PB169636BL	Decachlorobiphen	1	20	20.0	100		57	171
		Tetrachloro-m-xyl	1	20	19.1	96		61	148
		Decachlorobiphen	2	20	21.2	106		57	171
		Tetrachloro-m-xyl	2	20	20.7	103		61	148
PB169636BS	PB169636BS	Decachlorobiphen	1	20	18.4	92		57	171
		Tetrachloro-m-xyl	1	20	15.1	75		61	148
		Decachlorobiphen	2	20	19.8	99		57	171
		Tetrachloro-m-xyl	2	20	21.7	109		61	148
I.BLK-PD090232.D	PIBLK-PD090232.D	Decachlorobiphen	1	20	19.9	99		57	171
		Tetrachloro-m-xyl	1	20	17.1	86		61	148
		Decachlorobiphen	2	20	21.2	106		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148