



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3015</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Alliance Technical Group, LLC - Newark</u>	Datafile :	<u>PD090242.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169637BS (Column 1)	Toxaphene	2	2.00	ug/L	100				80	120	
PB169637BS (Column 2)	Toxaphene	2	2.30	ug/L	115				80	120	