

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Instrument ID: ECD_Q

Calibration Date(s): 10/07/2025 10/07/2025

Calibration Times: 08:57 11:09

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PQ070947.D</u>		CF 0.25 = <u>PQ070948.D</u>			
CF 0.1 = <u>PQ070949.D</u>		CF 0.05 = <u>PQ070950.D</u>		CF 0.025 = <u>PQ070951.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	222270000000	281868000000	274450000000	278220000000	279426000000	267247000000	9
EDB	86215800000	108510000000	101887000000	111775000000	97848200000	101247000000	10

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3015

Instrument ID: ECD_Q

Calibration Date(s): 10/07/2025 10/07/2025

Calibration Times: 08:57 11:09

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID: CF 0.5 = <u>PQ070947.D</u> CF 0.25 = <u>PQ070948.D</u> CF 0.1 = <u>PQ070949.D</u> CF 0.05 = <u>PQ070950.D</u> CF 0.025 = <u>PQ070951.D</u>							
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	240227000000	298226000000	301754000000	309057000000	308382000000	291529000000	10
EDB	88910100000	110714000000	102294000000	114762000000	104137000000	104164000000	10