



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8011

Client: Alliance Technical Group, LLC - Newark

Datafile : PQ070958.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170012BSD (Column 1)	DBCP	0.25	0.24	ug/L	96	4			70	130	20
	EDB	0.25	0.26	ug/L	104	4			70	130	20
PB170012BSD (Column 2)	DBCP	0.25	0.24	ug/L	96	4			70	130	20
	EDB	0.25	0.25	ug/L	100	4			70	130	20