

Surrogate Summary

SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169719BL	PB169719BL	2-Fluorophenol	150	119	79		23	138
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	73.4	73		67	132
		2-Fluorobiphenyl	100	75.1	75		52	132
		2,4,6-Tribromophenol	150	111	74		44	137
		Terphenyl-d14	100	71.7	72		42	152
PB169719BS	PB169719BS	2-Fluorophenol	150	116	77		23	138
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	69.0	69		67	132
		2-Fluorobiphenyl	100	68.9	69		52	132
		2,4,6-Tribromophenol	150	113	75		44	137
		Terphenyl-d14	100	72.6	73		42	152
PB169719BSD	PB169719BSD	2-Fluorophenol	150	116	78		23	138
		Phenol-d6	150	110	74		10	134
		Nitrobenzene-d5	100	69.8	70		67	132
		2-Fluorobiphenyl	100	69.9	70		52	132
		2,4,6-Tribromophenol	150	113	76		44	137
		Terphenyl-d14	100	72.7	73		42	152
Q3015-03	WP0925-PT-BN-WP	2-Fluorophenol	150	145	96		23	138
		Phenol-d6	150	139	93		10	134
		Nitrobenzene-d5	100	89.8	90		67	132
		2-Fluorobiphenyl	100	92.3	92		52	132
		2,4,6-Tribromophenol	150	150	100		44	137
		Terphenyl-d14	100	92.2	92		42	152
Q3015-03DL	WP0925-PT-BN-WPDL	2-Fluorophenol	150	130	87		23	138
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	73.8	74		67	132
		2-Fluorobiphenyl	100	84.7	85		52	132
		2,4,6-Tribromophenol	150	118	78		44	137
		Terphenyl-d14	100	88.0	88		42	152