

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
 Data File : BP025732.D
 Acq On : 11 Sep 2025 14:24
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Sep 11 16:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:16:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.761	152	156685	20.000	ng	0.00
21) Naphthalene-d8	10.531	136	657507	20.000	ng	0.00
39) Acenaphthene-d10	14.384	164	457251	20.000	ng	0.00
64) Phenanthrene-d10	17.189	188	963896	20.000	ng	0.00
76) Chrysene-d12	21.642	240	964091	20.000	ng	0.00
86) Perylene-d12	24.995	264	1072676	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.384	112	1519863	156.516	ng	0.00
7) Phenol-d6	6.949	99	2060064	159.607	ng	0.00
23) Nitrobenzene-d5	8.908	82	2312828	155.163	ng	0.00
42) 2,4,6-Tribromophenol	15.901	330	901504	158.067	ng	0.00
45) 2-Fluorobiphenyl	12.996	172	4820104	140.360	ng	0.00
79) Terphenyl-d14	19.913	244	7841805	146.328	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.308	88	293942	72.123	ng	100
3) Pyridine	3.708	79	864675	77.049	ng	97
4) n-Nitrosodimethylamine	3.614	42	406094	76.687	ng	100
6) Aniline	7.096	93	1418325	80.049	ng	99
8) 2-Chlorophenol	7.337	128	852861	80.059	ng	99
9) Benzaldehyde	6.914	77	614881	79.569	ng	100
10) Phenol	6.972	94	1090983	80.162	ng	99
11) bis(2-Chloroethyl)ether	7.190	93	855627	78.251	ng	98
12) 1,3-Dichlorobenzene	7.649	146	917305	75.659	ng	97
13) 1,4-Dichlorobenzene	7.796	146	935253	75.728	ng	99
14) 1,2-Dichlorobenzene	8.108	146	905715	75.489	ng	99
15) Benzyl Alcohol	8.002	79	865116	81.113	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.278	45	1365629	76.370	ng	99
17) 2-Methylphenol	8.208	107	747211	81.433	ng	100
18) Hexachloroethane	8.825	117	362679	76.881	ng	96
19) n-Nitroso-di-n-propyla...	8.560	70	721517	81.103	ng	99
20) 3+4-Methylphenols	8.531	107	1023340	79.634	ng	99
22) Acetophenone	8.572	105	1366991	77.798	ng	# 98
24) Nitrobenzene	8.949	77	1052990	78.746	ng	98
25) Isophorone	9.484	82	2070061	79.012	ng	100
26) 2-Nitrophenol	9.649	139	504228	84.931	ng	96
27) 2,4-Dimethylphenol	9.713	122	842469	80.887	ng	99
28) bis(2-Chloroethoxy)met...	9.937	93	1190837	78.631	ng	99
29) 2,4-Dichlorophenol	10.190	162	850329	81.883	ng	99
30) 1,2,4-Trichlorobenzene	10.390	180	902735	77.134	ng	99
31) Naphthalene	10.578	128	2657135	74.948	ng	99
32) Benzoic acid	9.925	122	680741	87.910	ng	98
33) 4-Chloroaniline	10.684	127	1237939	83.774	ng	98
34) Hexachlorobutadiene	10.866	225	571616	75.871	ng	100
35) Caprolactam	11.513	113	324639	81.266	ng	96
36) 4-Chloro-3-methylphenol	11.825	107	998570	80.654	ng	96
37) 2-Methylnaphthalene	12.190	142	1750889	76.891	ng	99
38) 1-Methylnaphthalene	12.407	142	1849786	76.178	ng	99
40) 1,2,4,5-Tetrachloroben...	12.554	216	1039713	75.038	ng	100
41) Hexachlorocyclopentadiene	12.537	237	641317	84.983	ng	98
43) 2,4,6-Trichlorophenol	12.807	196	725824	79.683	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
 Data File : BP025732.D
 Acq On : 11 Sep 2025 14:24
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Sep 11 16:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:16:10 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.884	196	806956	79.601	ng	97
46) 1,1'-Biphenyl	13.201	154	2497929	74.426	ng	100
47) 2-Chloronaphthalene	13.243	162	1962349	74.255	ng	98
48) 2-Nitroaniline	13.454	65	733091	83.218	ng	97
49) Acenaphthylene	14.107	152	3311359	75.625	ng	99
50) Dimethylphthalate	13.843	163	2615037	74.909	ng	99
51) 2,6-Dinitrotoluene	13.954	165	580821	78.530	ng	100
52) Acenaphthene	14.443	154	1846398	73.121	ng	100
53) 3-Nitroaniline	14.290	138	649230	83.474	ng	100
54) 2,4-Dinitrophenol	14.507	184	382175	79.506	ng	93
55) Dibenzofuran	14.790	168	2999483	72.882	ng	99
56) 4-Nitrophenol	14.625	139	527475	78.823	ng	98
57) 2,4-Dinitrotoluene	14.760	165	853243	81.062	ng	98
58) Fluorene	15.454	166	2363617	72.225	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.019	232	722846	79.155	ng	98
60) Diethylphthalate	15.225	149	2678679	75.646	ng	100
61) 4-Chlorophenyl-phenyle...	15.442	204	1190470	72.181	ng	97
62) 4-Nitroaniline	15.472	138	667122	83.790	ng	99
63) Azobenzene	15.742	77	2578205	75.299	ng	99
65) 4,6-Dinitro-2-methylph...	15.542	198	537070	83.783	ng	97
66) n-Nitrosodiphenylamine	15.666	169	2109969	71.530	ng	100
67) 4-Bromophenyl-phenylether	16.354	248	792238	74.816	ng	97
68) Hexachlorobenzene	16.478	284	892102	75.546	ng	96
69) Atrazine	16.648	200	876809	77.763	ng	99
70) Pentachlorophenol	16.837	266	629802	80.808	ng	98
71) Phenanthrene	17.236	178	3960390	72.431	ng	100
72) Anthracene	17.331	178	4027019	72.588	ng	100
73) Carbazole	17.607	167	3852545	74.694	ng	99
74) Di-n-butylphthalate	18.195	149	4705330	74.340	ng	100
75) Fluoranthene	19.325	202	4786922	72.998	ng	100
77) Benzidine	19.519	184	2065703	74.384	ng	100
78) Pyrene	19.701	202	4911699	76.353	ng	100
80) Butylbenzylphthalate	20.642	149	2268963	80.443	ng	98
81) Benzo(a)anthracene	21.619	228	5031313	76.229	ng	100
82) 3,3'-Dichlorobenzidine	21.536	252	1705366	74.725	ng	99
83) Chrysene	21.689	228	4667905	74.190	ng	99
84) Bis(2-ethylhexyl)phtha...	21.542	149	3154011	76.431	ng	100
85) Di-n-octyl phthalate	22.824	149	5799888	79.012	ng	100
87) Indeno(1,2,3-cd)pyrene	28.848	276	5948880	76.261	ng	# 84
88) Benzo(b)fluoranthene	23.936	252	5088293	77.570	ng	99
89) Benzo(k)fluoranthene	24.013	252	4938760	73.783	ng	99
90) Benzo(a)pyrene	24.842	252	4896947	76.231	ng	99
91) Dibenzo(a,h)anthracene	28.918	278	4903653	76.817	ng	98
92) Benzo(g,h,i)perylene	30.036	276	4826771	76.886	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
 Data File : BP025732.D
 Acq On : 11 Sep 2025 14:24
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Sep 11 16:26:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:16:10 2025
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.761 min Scan# 81
 Delta R.T. 0.001 min
 Lab File: BP025732.D
 Acq: 11 Sep 2025 14:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:152 Resp: 156685
 Ion Ratio Lower Upper
 152 100
 150 153.0 124.3 186.5
 115 61.0 50.3 75.5



#2
 1,4-Dioxane
 Concen: 72.123 ng
 RT: 3.308 min Scan# 63
 Delta R.T. 0.000 min
 Lab File: BP025732.D
 Acq: 11 Sep 2025 14:24

Tgt Ion: 88 Resp: 293942
 Ion Ratio Lower Upper
 88 100
 58 81.8 65.5 98.3
 43 31.6 25.2 37.8









#3
Pyridine
Concen: 77.049 ng
RT: 3.708 min Scan# 131
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion: 79 Resp: 864675
Ion Ratio Lower Upper
79 100
52 77.1 63.7 95.5
51 39.1 32.4 48.6



#4
n-Nitrosodimethylamine
Concen: 76.687 ng
RT: 3.614 min Scan# 115
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 42 Resp: 406094
Ion Ratio Lower Upper
42 100
74 123.8 99.2 148.8
44 6.8 6.4 9.6









#5
2-Fluorophenol
Concen: 156.516 ng
RT: 5.384 min Scan# 416
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:112 Resp: 1519863
Ion Ratio Lower Upper
112 100
64 62.6 50.2 75.4
63 33.0 26.7 40.1



#6
Aniline
Concen: 80.049 ng
RT: 7.096 min Scan# 707
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 93 Resp: 1418325
Ion Ratio Lower Upper
93 100
66 41.5 32.6 49.0
65 20.9 16.2 24.2









#7
Phenol-d6
Concen: 159.607 ng
RT: 6.949 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion: 99 Resp: 2060064
Ion Ratio Lower Upper
99 100
42 19.6 15.6 23.4
71 37.3 28.6 43.0



#8
2-Chlorophenol
Concen: 80.059 ng
RT: 7.337 min Scan# 748
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:128 Resp: 852861
Ion Ratio Lower Upper
128 100
130 32.6 12.6 52.6
64 50.2 31.6 71.6









#9
Benzaldehyde
Concen: 79.569 ng
RT: 6.914 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion: 77 Resp: 614881
Ion Ratio Lower Upper
77 100
105 95.0 75.2 115.2
106 91.3 71.3 111.3



#10
Phenol
Concen: 80.162 ng
RT: 6.972 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 94 Resp: 1090983
Ion Ratio Lower Upper
94 100
65 30.5 10.2 50.2
66 42.3 21.4 61.4









#11
bis(2-Chloroethyl)ether
Concen: 78.251 ng
RT: 7.190 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion: 93 Resp: 855627
Ion Ratio Lower Upper
93 100
63 80.8 62.6 102.6
95 32.5 13.2 53.2



#12
1,3-Dichlorobenzene
Concen: 75.659 ng
RT: 7.649 min Scan# 801
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:146 Resp: 917305
Ion Ratio Lower Upper
146 100
148 63.3 52.7 79.1
75 34.8 26.4 39.6









#13
1,4-Dichlorobenzene
Concen: 75.728 ng
RT: 7.796 min Scan# 81
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:146 Resp: 935253
Ion Ratio Lower Upper
146 100
148 64.0 52.0 78.0
111 42.6 33.9 50.9



#14
1,2-Dichlorobenzene
Concen: 75.489 ng
RT: 8.108 min Scan# 879
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:146 Resp: 905715
Ion Ratio Lower Upper
146 100
148 63.9 51.7 77.5
111 44.9 35.5 53.3









#15
Benzyl Alcohol
Concen: 81.113 ng
RT: 8.002 min Scan# 861
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 865116
Ion Ratio Lower Upper
79 100
108 73.1 58.7 88.1
77 64.5 50.2 75.2



#16
2,2'-oxybis(1-Chloropropane)
Concen: 76.370 ng
RT: 8.278 min Scan# 908
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 45 Resp: 1365629
Ion Ratio Lower Upper
45 100
77 13.4 0.0 33.1
79 10.3 0.0 30.0









#17
2-Methylphenol
Concen: 81.433 ng
RT: 8.208 min Scan# 895
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:	107	Resp:	747211
Ion	Ratio	Lower	Upper
107	100		
108	112.1	89.6	134.4
77	49.3	39.3	58.9
79	48.1	37.8	56.6



#18
Hexachloroethane
Concen: 76.881 ng
RT: 8.825 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:	117	Resp:	362679
Ion	Ratio	Lower	Upper
117	100		
119	94.3	77.1	115.7
201	101.0	85.6	128.4





#19

n-Nitroso-di-n-propylamine

Concen: 81.103 ng

RT: 8.560 min Scan# 91

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080

Tgt Ion: 70 Resp: 721517





Ion Ratio Lower Upper

70 100

42 60.8 49.0 73.4

101 10.0 7.9 11.9

130 20.5 16.6 25.0



#20

3+4-Methylphenols

Concen: 79.634 ng

RT: 8.531 min Scan# 951

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Tgt Ion: 107 Resp: 1023340




Ion Ratio Lower Upper


107 100

108 89.5 67.8 107.8

77 35.7 16.1 56.1

79 30.5 10.5 50.5









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.531 min Scan# 1291
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 657507




Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	9.1	7.5	11.3
68	6.5	5.0	7.6




#22
Acetophenone
Concen: 77.798 ng
RT: 8.572 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:105 Resp: 1366991

Ion	Ratio	Lower	Upper
105	100		
71	4.8	1.8	2.8
51	30.6	23.5	35.3
120	22.0	17.7	26.5





#23

Nitrobenzene-d5

Concen: 155.163 ng

RT: 8.908 min Scan# 1015

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion: 82 Resp: 2312828

Ion Ratio Lower Upper

82 100

128 41.9 32.9 49.3

54 54.5 43.4 65.2



#24

Nitrobenzene

Concen: 78.746 ng

RT: 8.949 min Scan# 1022

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion: 77 Resp: 1052990


Ion Ratio Lower Upper

77 100

123 44.4 34.4 51.6

65 14.3 10.8 16.2








#25
Isophorone
Concen: 79.012 ng
RT: 9.484 min Scan# 1113
Delta R.T. 0.018 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion: 82 Resp: 2070061
Ion Ratio Lower Upper
82 100
95 7.2 5.8 8.8
138 17.0 13.5 20.3



#26
2-Nitrophenol
Concen: 84.931 ng
RT: 9.649 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 139 Resp: 504228
Ion Ratio Lower Upper
139 100
109 34.0 29.0 43.4
65 55.0 46.9 70.3





#27

2,4-Dimethylphenol

Concen: 80.887 ng

RT: 9.713 min Scan# 1152

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:122 Resp: 842469

Ion Ratio Lower Upper

122 100

107 123.6 97.9 146.9

121 58.7 47.4 71.0



#28

bis(2-Chloroethoxy)methane

Concen: 78.631 ng

RT: 9.937 min Scan# 1190

Delta R.T. -0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion: 93 Resp: 1190837


Ion Ratio Lower Upper

93 100

95 32.6 25.6 38.4

123 11.0 9.2 13.8





#29

2,4-Dichlorophenol

Concen: 81.883 ng

RT: 10.190 min Scan# 1232

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:162 Resp: 850329

Ion Ratio Lower Upper

162 100

164 64.3 44.7 84.7

98 37.5 18.1 58.1



#30

1,2,4-Trichlorobenzene

Concen: 77.134 ng

RT: 10.390 min Scan# 1267

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:180 Resp: 902735


Ion Ratio Lower Upper

180 100

182 95.6 76.0 114.0

145 30.3 24.7 37.1





#31

Naphthalene

Concen: 74.948 ng

RT: 10.578 min Scan# 1299

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:128 Resp: 2657135

Ion Ratio Lower Upper

128 100

129 11.0 8.6 12.8

127 13.2 10.2 15.4



#32

Benzoic acid

Concen: 87.910 ng

RT: 9.925 min Scan# 1188

Delta R.T. 0.047 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:122 Resp: 680741


Ion Ratio Lower Upper

122 100

105 129.3 111.5 151.5

77 100.2 81.8 121.8









#33
4-Chloroaniline
Concen: 83.774 ng
RT: 10.684 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:127 Resp: 1237939




Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.1	37.7
65	34.3	26.4	39.6
92	20.8	15.9	23.9




#34
Hexachlorobutadiene
Concen: 75.871 ng
RT: 10.866 min Scan# 1348
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:225 Resp: 571616

Ion	Ratio	Lower	Upper
225	100		
223	62.4	50.1	75.1
227	63.5	51.2	76.8









#35
Caprolactam
Concen: 81.266 ng
RT: 11.513 min Scan# 1458
Delta R.T. 0.029 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:113 Resp: 324639
Ion Ratio Lower Upper
113 100
55 180.6 165.5 205.5
56 135.9 121.1 161.1



#36
4-Chloro-3-methylphenol
Concen: 80.654 ng
RT: 11.825 min Scan# 1511
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:107 Resp: 998570
Ion Ratio Lower Upper
107 100
144 24.3 19.4 29.2
142 75.3 63.8 95.6









#37
2-Methylnaphthalene
Concen: 76.891 ng
RT: 12.190 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:142 Resp: 1750889
Ion Ratio Lower Upper
142 100
141 88.5 69.6 104.4
115 37.5 30.0 45.0



#38
1-Methylnaphthalene
Concen: 76.178 ng
RT: 12.407 min Scan# 1610
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:142 Resp: 1849786
Ion Ratio Lower Upper
142 100
141 91.6 72.7 109.1
116 4.3 3.5 5.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.384 min Scan# 1946

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:164 Resp: 457251

Ion Ratio Lower Upper

164 100

162 100.9 80.6 120.8

160 44.9 35.0 52.6



#40

1,2,4,5-Tetrachlorobenzene

Concen: 75.038 ng

RT: 12.554 min Scan# 1635

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Tgt Ion:216 Resp: 1039713




Ion Ratio Lower Upper


216 100

214 78.4 62.4 93.6

179 21.1 17.0 25.6

108 18.9 15.2 22.8









#41
Hexachlorocyclopentadiene
Concen: 84.983 ng
RT: 12.537 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:	237	Resp:	641317
Ion	Ratio	Lower	Upper
237	100		
235	63.6	41.9	81.9
272	12.3	0.0	31.8



#42
2,4,6-Tribromophenol
Concen: 158.067 ng
RT: 15.901 min Scan# 2204
Delta R.T. -0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:	330	Resp:	901504
Ion	Ratio	Lower	Upper
330	100		
332	96.7	76.9	115.3
141	37.7	31.0	46.4









#43
2,4,6-Trichlorophenol
Concen: 79.683 ng
RT: 12.807 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:196 Resp: 725824
Ion Ratio Lower Upper
196 100
198 96.1 77.2 115.8
200 31.7 24.2 36.2



#44
2,4,5-Trichlorophenol
Concen: 79.601 ng
RT: 12.884 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:196 Resp: 806956
Ion Ratio Lower Upper
196 100
198 95.5 77.8 116.8
97 50.6 43.3 64.9
132 27.1 22.3 33.5









#45
2-Fluorobiphenyl
Concen: 140.360 ng
RT: 12.996 min Scan# 1710
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:172 Resp: 4820104
Ion Ratio Lower Upper
172 100
171 35.2 27.8 41.6
170 23.6 18.6 27.8



#46
1,1'-Biphenyl
Concen: 74.426 ng
RT: 13.201 min Scan# 1745
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:154 Resp: 2497929
Ion Ratio Lower Upper
154 100
153 40.1 19.7 59.7
76 13.4 0.0 33.3









#47
2-Chloronaphthalene
Concen: 74.255 ng
RT: 13.243 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:162 Resp: 1962349
Ion Ratio Lower Upper
162 100
127 41.7 32.2 48.4
164 32.4 26.5 39.7



#48
2-Nitroaniline
Concen: 83.218 ng
RT: 13.454 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion: 65 Resp: 733091
Ion Ratio Lower Upper
65 100
92 67.4 53.1 79.7
138 101.1 77.8 116.8





#49

Acenaphthylene

Concen: 75.625 ng

RT: 14.107 min Scan# 1898

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:152 Resp: 3311359

Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.6

153 13.0 10.5 15.7



#50

Dimethylphthalate

Concen: 74.909 ng

RT: 13.843 min Scan# 1854

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:163 Resp: 2615037


Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.2

164 10.4 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 78.530 ng
RT: 13.954 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:165 Resp: 580821
Ion Ratio Lower Upper
165 100
63 59.9 48.1 72.1
89 57.1 45.4 68.0



#52
Acenaphthene
Concen: 73.121 ng
RT: 14.443 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:154 Resp: 1846398
Ion Ratio Lower Upper
154 100
153 111.4 89.4 134.0
152 52.8 42.5 63.7









#53
3-Nitroaniline
Concen: 83.474 ng
RT: 14.290 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:138 Resp: 649230
Ion Ratio Lower Upper
138 100
108 11.3 8.7 13.1
92 118.4 94.6 142.0



#54
2,4-Dinitrophenol
Concen: 79.506 ng
RT: 14.507 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:184 Resp: 382175
Ion Ratio Lower Upper
184 100
63 72.4 52.5 78.7
154 68.1 51.5 77.3





#55

Dibenzofuran

Concen: 72.882 ng

RT: 14.790 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:168 Resp: 2999483

Ion Ratio Lower Upper

168 100

139 39.0 31.7 47.5

169 13.5 10.8 16.2



#56

4-Nitrophenol

Concen: 78.823 ng

RT: 14.625 min Scan# 1987

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:139 Resp: 527475


Ion Ratio Lower Upper

139 100

109 84.0 62.9 102.9

65 105.6 83.4 123.4





#57

2,4-Dinitrotoluene

Concen: 81.062 ng

RT: 14.760 min Scan# 2009

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080

Tgt Ion:165 Resp: 853243





Ion Ratio Lower Upper

165 100

63 49.2 37.7 56.5

89 75.4 61.0 91.6

182 2.9 2.4 3.6



#58

Fluorene

Concen: 72.225 ng

RT: 15.454 min Scan# 2128

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:166 Resp: 2363617


Ion Ratio Lower Upper

166 100

165 97.0 77.7 116.5

167 13.2 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 79.155 ng
RT: 15.019 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:	232	Resp:	722846
Ion	Ratio	Lower	Upper
232	100		
131	44.8	36.7	55.1
130	2.4	2.0	3.0
166	31.0	25.4	38.2



#60
Diethylphthalate
Concen: 75.646 ng
RT: 15.225 min Scan# 2089
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:	149	Resp:	2678679
Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.7	26.5
150	12.7	10.0	15.0









#61
4-Chlorophenyl-phenylether
Concen: 72.181 ng
RT: 15.442 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:204 Resp: 1190470
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 59.2 49.7 74.5



#62
4-Nitroaniline
Concen: 83.790 ng
RT: 15.472 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:138 Resp: 667122
Ion Ratio Lower Upper
138 100
92 53.4 33.3 73.3
108 87.6 66.7 106.7





#63

Azobenzene

Concen: 75.299 ng

RT: 15.742 min Scan# 2176

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080

Tgt Ion: 77 Resp: 2578205





Ion Ratio Lower Upper

77 100

182 26.0 5.2 45.2

105 21.2 1.1 41.1

51 31.8 11.6 51.6



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.189 min Scan# 2423

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:188 Resp: 963896


Ion Ratio Lower Upper

188 100

94 8.7 7.4 11.0

80 10.4 7.7 11.5









#65
4,6-Dinitro-2-methylphenol
Concen: 83.783 ng
RT: 15.542 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:198 Resp: 537070
Ion Ratio Lower Upper
198 100
51 50.6 33.4 73.4
105 43.2 24.3 64.3



#66
n-Nitrosodiphenylamine
Concen: 71.530 ng
RT: 15.666 min Scan# 2164
Delta R.T. 0.012 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:169 Resp: 2109969
Ion Ratio Lower Upper
169 100
168 65.6 52.6 78.8
167 36.2 28.5 42.7









#67
4-Bromophenyl-phenylether
Concen: 74.816 ng
RT: 16.354 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:248 Resp: 792238
Ion Ratio Lower Upper
248 100
250 97.8 77.5 116.3
141 80.5 60.2 90.4



#68
Hexachlorobenzene
Concen: 75.546 ng
RT: 16.478 min Scan# 2302
Delta R.T. 0.006 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:284 Resp: 892102
Ion Ratio Lower Upper
284 100
142 35.8 31.3 46.9
249 32.3 27.3 40.9





#69

Atrazine

Concen: 77.763 ng

RT: 16.648 min Scan# 2331

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:200 Resp: 876809

Ion Ratio Lower Upper

200 100

173 24.9 4.2 44.2

215 47.8 27.3 67.3



#70

Pentachlorophenol

Concen: 80.808 ng

RT: 16.837 min Scan# 2363

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:266 Resp: 629802


Ion Ratio Lower Upper

266 100

268 64.2 49.8 74.8

264 63.6 50.2 75.2





#71

Phenanthrene

Concen: 72.431 ng

RT: 17.236 min Scan# 24

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:178 Resp: 3960390

Ion Ratio Lower Upper

178 100

176 19.4 15.4 23.2

179 15.5 12.2 18.4



#72

Anthracene

Concen: 72.588 ng

RT: 17.331 min Scan# 2447

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:178 Resp: 4027019


Ion Ratio Lower Upper

178 100

176 18.7 14.8 22.2

179 15.5 12.2 18.4





#73

Carbazole

Concen: 74.694 ng

RT: 17.607 min Scan# 2493

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :





BNA_P

ClientSampleId :

SSTDICC080

Tgt Ion:167 Resp: 3852545

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.0	25.4
139	13.2	10.6	16.0



#74

Di-n-butylphthalate

Concen: 74.340 ng

RT: 18.195 min Scan# 2594




Delta R.T. 0.012 min


Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Tgt Ion:149 Resp: 4705330

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.0
104	5.4	4.3	6.5





#75

Fluoranthene

Concen: 72.998 ng

RT: 19.325 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:202 Resp: 4786922

Ion Ratio Lower Upper

202 100

101 9.9 0.0 30.4

203 17.4 0.0 37.4



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.642 min Scan# 3180

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:240 Resp: 964091


Ion Ratio Lower Upper

240 100

120 9.7 8.0 12.0

236 26.1 20.7 31.1





#77

Benzidine

Concen: 74.384 ng

RT: 19.519 min Scan# 2819

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:184 Resp: 2065703

Ion Ratio Lower Upper

184 100

185 14.1 11.5 17.3

183 11.5 9.2 13.8



#78

Pyrene

Concen: 76.353 ng

RT: 19.701 min Scan# 2850

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:202 Resp: 4911699


Ion Ratio Lower Upper

202 100

200 21.1 17.0 25.6

203 17.4 13.9 20.9





#79

Terphenyl-d14

Concen: 146.328 ng

RT: 19.913 min Scan# 2886

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:244 Resp: 7841805

Ion Ratio Lower Upper

244 100

212 7.5 5.9 8.9

122 8.4 7.1 10.7



#80

Butylbenzylphthalate

Concen: 80.443 ng

RT: 20.642 min Scan# 3010

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:149 Resp: 2268963


Ion Ratio Lower Upper

149 100

91 73.3 56.6 85.0

206 21.4 17.4 26.0





#81

Benzo(a)anthracene

Concen: 76.229 ng

RT: 21.619 min Scan# 3176

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:228 Resp: 5031313

Ion Ratio Lower Upper

228 100

226 27.1 21.6 32.4

229 19.9 15.7 23.5



#82

3,3'-Dichlorobenzidine

Concen: 74.725 ng

RT: 21.536 min Scan# 3162

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:252 Resp: 1705366


Ion Ratio Lower Upper

252 100

254 65.1 51.4 77.0

126 9.5 8.3 12.5





#83

Chrysene

Concen: 74.190 ng

RT: 21.689 min Scan# 3188

Delta R.T. 0.000 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:228 Resp: 4667905

Ion Ratio Lower Upper

228 100

226 30.2 23.8 35.6

229 19.7 15.6 23.4



#84

Bis(2-ethylhexyl)phthalate

Concen: 76.431 ng

RT: 21.542 min Scan# 3163

Delta R.T. -0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:149 Resp: 3154011


Ion Ratio Lower Upper

149 100

167 28.3 22.9 34.3

279 5.0 3.8 5.8





#85

Di-n-octyl phthalate

Concen: 79.012 ng

RT: 22.824 min Scan# 3382

Delta R.T. -0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:149 Resp: 5799888

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9

43 9.8 8.0 12.0



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.995 min Scan# 3750

Delta R.T. -0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:264 Resp: 1072676


Ion Ratio Lower Upper

264 100

260 23.3 18.3 27.5

265 21.4 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 76.261 ng

RT: 28.848 min Scan# 4403

Delta R.T. 0.012 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24

Instrument :

BNA_P

ClientSampleId :

SSTDICC080





Tgt Ion:276 Resp: 5948880

Ion Ratio Lower Upper

276 100

138 15.6 13.4 20.2

277 25.4 9.1 13.7#



#88

Benzo(b)fluoranthene

Concen: 77.570 ng

RT: 23.936 min Scan# 3570

Delta R.T. 0.006 min

Lab File: BP025732.D

Acq: 11 Sep 2025 14:24




Tgt Ion:252 Resp: 5088293


Ion Ratio Lower Upper

252 100

253 22.0 17.9 26.9

125 9.5 7.3 10.9









#89
Benzo(k)fluoranthene
Concen: 73.783 ng
RT: 24.013 min Scan# 3581
Delta R.T. 0.012 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24




Instrument :
BNA_P
ClientSampleId :
SSTDICC080


Tgt Ion:252 Resp: 4938760
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 8.9 7.5 11.3



#90
Benzo(a)pyrene
Concen: 76.231 ng
RT: 24.842 min Scan# 3724
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:252 Resp: 4896947
Ion Ratio Lower Upper
252 100
253 21.7 17.2 25.8
125 9.5 8.3 12.5









#91
Dibenzo(a,h)anthracene
Concen: 76.817 ng
RT: 28.918 min Scan# 4417
Delta R.T. 0.000 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:278 Resp: 4903653
Ion Ratio Lower Upper
278 100
139 13.9 12.1 18.1
279 23.7 19.6 29.4



#92
Benzo(g,h,i)perylene
Concen: 76.886 ng
RT: 30.036 min Scan# 4607
Delta R.T. 0.024 min
Lab File: BP025732.D
Acq: 11 Sep 2025 14:24

Tgt Ion:276 Resp: 4826771
Ion Ratio Lower Upper
276 100
277 24.2 18.7 28.1
138 20.5 16.0 24.0

