

Surrogate Summary

SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169719BL	PB169719BL	2-Fluorophenol	150	119	79		23	138
		Phenol-d6	150	109	73		10	134
		2,4,6-Tribromophenol	150	111	74		44	137
PB169719BS	PB169719BS	2-Fluorophenol	150	116	77		23	138
		Phenol-d6	150	109	73		10	134
		2,4,6-Tribromophenol	150	113	75		44	137
PB169719BSD	PB169719BSD	2-Fluorophenol	150	116	78		23	138
		Phenol-d6	150	110	74		10	134
		2,4,6-Tribromophenol	150	113	76		44	137
Q3015-06	WP0925-PT-ACIDS-WP	2-Fluorophenol	150	145	97		23	138
		Phenol-d6	150	142	95		10	134
		2,4,6-Tribromophenol	150	141	94		44	137
Q3015-06DL	WP0925-PT-ACIDS-WPDL	2-Fluorophenol	150	143	96		23	138
		Phenol-d6	150	132	88		10	134
		2,4,6-Tribromophenol	150	135	90		44	137