

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3015

Client ID : SSTDCCC040

Date Analyzed: 09/18/2025

Lab File ID: BP025779.D

Time Analyzed: 10:15

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	158404	7.754	625211	10.52	407763	14.38
UPPER LIMIT	316808	8.254	1250420	11.019	815526	14.877
LOWER LIMIT	79202	7.254	312606	10.019	203882	13.877
EPA SAMPLE NO.						
01 WP0925-PT-ACIDS-WP	132437	7.70	521061	10.48	332877	14.36
02 WP0925-PT-ACIDS-WPDL	153740	7.71	591198	10.50	370117	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.