

Surrogate Summary

SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169712BL	PB169712BL	2-Methylnaphthalene-d10	0.4	0.32	81		20	139
		Fluoranthene-d10	0.4	0.36	89		54	157
		Nitrobenzene-d5	0.4	0.38	95		27	154
		2-Fluorobiphenyl	0.4	0.39	97		30	155
		Terphenyl-d14	0.4	0.41	101		54	175
PB169712BS	PB169712BS	2-Methylnaphthalene-d10	0.4	0.35	88		20	139
		Fluoranthene-d10	0.4	0.34	86		54	157
		Nitrobenzene-d5	0.4	0.38	95		27	154
		2-Fluorobiphenyl	0.4	0.39	98		30	155
		Terphenyl-d14	0.4	0.40	100		54	175
PB169712BSD	PB169712BSD	2-Methylnaphthalene-d10	0.4	0.35	87		20	139
		Fluoranthene-d10	0.4	0.32	81		54	157
		Nitrobenzene-d5	0.4	0.39	97		27	154
		2-Fluorobiphenyl	0.4	0.40	99		30	155
		Terphenyl-d14	0.4	0.43	108		54	175
Q3015-05	WP0925-PT-BN-WP	2-Methylnaphthalene-d10	0.4	0	0	*	20	139
		Fluoranthene-d10	0.4	0	0	*	54	157
		Nitrobenzene-d5	0.4	110	27393	*	27	154
		2-Fluorobiphenyl	0.4	98.7	24663	*	30	155
		Terphenyl-d14	0.4	105	26136	*	54	175
Q3015-05DL	WP0925-PT-BN-WPDL	2-Methylnaphthalene-d10	0.4	0	0	*	20	139
		Fluoranthene-d10	0.4	0	0	*	54	157
		Nitrobenzene-d5	0.4	93.8	23450	*	27	154
		2-Fluorobiphenyl	0.4	99.8	24950	*	30	155
		Terphenyl-d14	0.4	114	28500	*	54	175