

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8270-Modified

Client: Alliance Technical Group, LLC - Newark

DataFile: BN037885.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169712BS	n-Nitrosodimethylamine	0.4	0.36	ug/L	90				33	130	
	Hexachlorobutadiene	0.4	0.37	ug/L	93				33	114	
	Hexachlorobenzene	0.4	0.37	ug/L	93				62	127	
	Atrazine	0.4	0.32	ug/L	80				20	150	
	1,4-Dioxane	0.4	0.32	ug/L	80				65	116	