

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8270-Modified

Client: Alliance Technical Group, LLC - Newark

DataFile: BN037888.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB169712BSD	n-Nitrosodimethylamine	0.4	0.40	ug/L	100	19			33	130	20
	Hexachlorobutadiene	0.4	0.39	ug/L	98	3			33	114	20
	Hexachlorobenzene	0.4	0.37	ug/L	93	0			62	127	20
	Atrazine	0.4	0.31	ug/L	78	10			20	150	20
	1,4-Dioxane	0.4	0.33	ug/L	83	6			65	116	27