

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3015

Client ID : SSTDCCC0.4

Date Analyzed: 09/24/2025

Lab File ID: BN037871.D

Time Analyzed: 09:54

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6027	7.818	16875	10.63	8790	14.47
UPPER LIMIT	12054	8.318	33750	11.126	17580	14.973
LOWER LIMIT	3013.5	7.318	8437.5	10.126	4395	13.973
EPA SAMPLE NO.						
01 PB169712BL	6892	7.83	18348	10.63	8435	14.47
02 WP0925-PT-BN-WP	6052	7.83	18152	10.63	9840	14.47
03 WP0925-PT-BN-WPDL	6044	7.83	17967	10.64	9123	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.