

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB169712BSD		SDG No.:	Q3015
Lab Sample ID:	PB169712BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group4
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037888.D	1	09/16/25 09:39	09/25/25 14:34	PB169712

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
534-52-1	4,6-Dinitro-2-methylphenol	0.39		0.12	0.20	ug/L
87-86-5	Pentachlorophenol	0.46		0.080	0.20	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	0.37		10 - 110	92%	SPK: 0.4
13127-88-3	Phenol-d6	0.36		10 - 110	89%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.31		20 - 159	76%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	6840	7.826			
1146-65-2	Naphthalene-d8	18500	10.626			
15067-26-2	Acenaphthene-d10	8850	14.473			
1517-22-2	Phenanthrene-d10	17000	17.223			
1719-03-5	Chrysene-d12	11900	21.402			
1520-96-3	Perylene-d12	9040	23.736			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products