

Surrogate Summary

SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169712BL	PB169712BL	2-Fluorophenol	0.4	0.36	89		10	110
		Phenol-d6	0.4	0.35	87		10	110
		2,4,6-Tribromophenol	0.4	0.13	32		20	159
PB169712BS	PB169712BS	2-Fluorophenol	0.4	0.37	92		10	110
		Phenol-d6	0.4	0.36	91		10	110
		2,4,6-Tribromophenol	0.4	0.31	78		20	159
PB169712BSD	PB169712BSD	2-Fluorophenol	0.4	0.37	92		10	110
		Phenol-d6	0.4	0.36	89		10	110
		2,4,6-Tribromophenol	0.4	0.31	76		20	159
Q3015-08	WP0925-PT-ACIDS-WP	2-Fluorophenol	150	186	124	*	10	110
		Phenol-d6	150	197	131	*	10	110
		2,4,6-Tribromophenol	150	319	213	*	20	159
Q3015-08DL	WP0925-PT-ACIDS-WPDL	2-Fluorophenol	150	166	111	*	10	110
		Phenol-d6	150	170	113	*	10	110
		2,4,6-Tribromophenol	150	173	115		20	159