

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB169712BL		SDG No.:	Q3015
Lab Sample ID:	PB169712BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group5
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037872.D	1	09/16/25 09:39	09/24/25 10:32	PB169712

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	0.040	U	0.040	0.10	ug/L
91-57-6	2-Methylnaphthalene	0.040	U	0.040	0.10	ug/L
208-96-8	Acenaphthylene	0.040	U	0.040	0.10	ug/L
83-32-9	Acenaphthene	0.040	U	0.040	0.10	ug/L
86-73-7	Fluorene	0.040	U	0.040	0.10	ug/L
85-01-8	Phenanthrene	0.030	U	0.030	0.10	ug/L
120-12-7	Anthracene	0.040	U	0.040	0.10	ug/L
206-44-0	Fluoranthene	0.030	U	0.030	0.10	ug/L
129-00-0	Pyrene	0.030	U	0.030	0.10	ug/L
56-55-3	Benzo(a)anthracene	0.040	U	0.040	0.10	ug/L
218-01-9	Chrysene	0.040	U	0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.040	U	0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.050	U	0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.040	U	0.040	0.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.050	U	0.050	0.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.050	U	0.050	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.32		20 - 139	81%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		54 - 157	89%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.38		27 - 154	95%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		30 - 155	97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		54 - 175	101%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	6890	7.826			
1146-65-2	Naphthalene-d8	18300	10.626			
15067-26-2	Acenaphthene-d10	8440	14.473			
1517-22-2	Phenanthrene-d10	16600	17.223			
1719-03-5	Chrysene-d12	13500	21.402			

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB169712BL		SDG No.:	Q3015
Lab Sample ID:	PB169712BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group5
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037872.D	1	09/16/25 09:39	09/24/25 10:32	PB169712

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1520-96-3	Perylene-d12	12200	23.733			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products