

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3015

Analytical Method: 8270-Modified

Client: Alliance Technical Group, LLC - Newark

DataFile: BN037888.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169712BSD	Naphthalene	0.4	0.36	ug/L	90	3			67	120	20
	2-Methylnaphthalene	0.4	0.31	ug/L	78	3			50	122	20
	Acenaphthylene	0.4	0.39	ug/L	98	0			60	119	20
	Acenaphthene	0.4	0.35	ug/L	88	3			65	119	20
	Fluorene	0.4	0.36	ug/L	90	3			58	122	20
	Phenanthrene	0.4	0.36	ug/L	90	3			65	117	20
	Anthracene	0.4	0.37	ug/L	93	6			57	118	20
	Fluoranthene	0.4	0.32	ug/L	80	10			53	126	20
	Pyrene	0.4	0.42	ug/L	105	2			54	124	20
	Benzo(a)anthracene	0.4	0.37	ug/L	93	0			54	130	20
	Chrysene	0.4	0.39	ug/L	98	5			64	126	20
	Benzo(b)fluoranthene	0.4	0.39	ug/L	98	3			65	121	20
	Benzo(k)fluoranthene	0.4	0.38	ug/L	95	5			72	119	20
	Benzo(a)pyrene	0.4	0.39	ug/L	98	3			68	120	20
	Indeno(1,2,3-cd)pyrene	0.4	0.35	ug/L	88	0			70	127	20
	Dibenz(a,h)anthracene	0.4	0.36	ug/L	90	3			65	121	20
	Benzo(g,h,i)perylene	0.4	0.37	ug/L	93	0			76	117	20