

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG No.: Q3015
 Instrument ID: BNA_P Calibration Date/Time: 09/17/2025 12:03
 Lab File ID: BP025769.D Init. Calib. Date(s): 09/11/2025 09/11/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 08:56 15:20
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.240	1.323		6.7	
Phenol-d6	1.648	1.712		3.9	
Nitrobenzene-d5	0.453	0.476		5.1	
2-Fluorobiphenyl	1.502	1.618		7.7	
2,4,6-Tribromophenol	0.249	0.276		10.8	
Atrazine	0.234	0.248		6.0	
Terphenyl-d14	1.112	1.118		0.5	

All other compounds must meet a minimum RRF of 0.010.