

Hit Summary Sheet
SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	WP0925-PT-VOA-WP							
Q3015-01	WP0925-PT-VOA- ¹ Water		Chloromethane	1.10	J	0.32	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Vinyl Chloride	71.0		0.26	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Bromomethane	56.2		1.40	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,1-Dichloroethene	88.1		0.23	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Acetone	140		1.50	25.0	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Methyl tert-butyl Ether	96.5		0.16	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Methylene Chloride	74.5		0.28	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		trans-1,2-Dichloroethene	88.2		0.23	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		2-Butanone	160		0.98	25.0	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		cis-1,2-Dichloroethene	120		0.19	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Chloroform	97.5		0.25	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Benzene	110		0.15	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,2-Dichloroethane	35.0		0.22	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Trichloroethene	90.4		0.090	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Dibromomethane	120		0.25	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Bromodichloromethane	17.1		0.22	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Toluene	35.4		0.14	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		t-1,3-Dichloropropene	62.0		0.17	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		cis-1,3-Dichloropropene	99.3		0.16	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,1,2-Trichloroethane	100		0.21	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		2-Hexanone	110		0.89	25.0	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Dibromochloromethane	68.9		0.18	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,2-Dibromoethane	33.3		0.15	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Tetrachloroethene	74.0		0.23	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Chlorobenzene	23.1		0.12	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,1,1,2-Tetrachloroethane	53.1		0.19	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Ethyl Benzene	48.6		0.13	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Total Xylenes	250		0.36	15.0	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		m/p-Xylenes	110		0.24	10.0	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		o-Xylene	140		0.12	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Bromoform	17.8		0.19	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		Isopropylbenzene	0.21	J	0.12	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,1,2,2-Tetrachloroethane	71.4		0.26	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		1,2,3-Trichloropropane	99.7		0.35	5.00	ug/L
Q3015-01	WP0925-PT-VOA- ¹ Water		sec-Butylbenzene	0.28	J	0.13	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3015-01	WP0925-PT-VOA-1	Water	1,3-Dichlorobenzene	0.22	J	0.16	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	1,4-Dichlorobenzene	110		0.19	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	n-Butylbenzene	0.43	J	0.15	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	1,2,4-Trichlorobenzene	0.61	J	0.20	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	Hexachlorobutadiene	1.40	J	0.30	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	Naphthalene	78.4		0.20	5.00	ug/L
Q3015-01	WP0925-PT-VOA-1	Water	1,2,3-Trichlorobenzene	0.59	J	0.20	5.00	ug/L
Total Voc :				2700				
Total Concentration:				2700				