

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047590.D
 Acq On : 16 Sep 2025 12:01
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Sep 22 01:52:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	153177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	274804	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	241818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	114743	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	396627	162.673	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 325.340%#	
35) Dibromofluoromethane	5.373	113	303728	164.802	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 329.600%#	
50) Toluene-d8	8.635	98	1041079	163.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 326.500%#	
62) 4-Bromofluorobenzene	11.067	95	399386	165.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 330.040%#	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.173	85	226166	142.586	ug/l	99
3) Chloromethane	1.301	50	297259	142.593	ug/l	99
4) Vinyl Chloride	1.380	62	302303	148.140	ug/l	98
5) Bromomethane	1.605	94	147541	114.019	ug/l	98
6) Chloroethane	1.691	64	195927	143.652	ug/l	99
7) Trichlorofluoromethane	1.892	101	453116	149.545	ug/l	99
8) Diethyl Ether	2.136	74	188006	151.684	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	267995	151.264	ug/l	99
10) Methyl Iodide	2.453	142	407242	147.384	ug/l	99
11) Tert butyl alcohol	2.959	59	215927	795.166	ug/l	100
12) 1,1-Dichloroethene	2.325	96	273692	150.388	ug/l	97
13) Acrolein	2.239	56	223763	894.203	ug/l	98
14) Allyl chloride	2.666	41	554779	147.727	ug/l	99
15) Acrylonitrile	3.063	53	790790	786.201	ug/l	99
16) Acetone	2.380	43	686119	647.201	ug/l	98
17) Carbon Disulfide	2.514	76	714503	143.412	ug/l	99
18) Methyl Acetate	2.703	43	404736	171.548	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	1015295	152.826	ug/l	100
20) Methylene Chloride	2.788	84	324032	144.398	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	291616	148.833	ug/l	98
22) Diisopropyl ether	3.757	45	1095726	150.551	ug/l	97
23) Vinyl Acetate	3.715	43	4504849	773.364	ug/l	100
24) 1,1-Dichloroethane	3.605	63	587129	151.776	ug/l	98
25) 2-Butanone	4.544	43	973938	736.502	ug/l	99
26) 2,2-Dichloropropane	4.471	77	486890	151.414	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	366262	152.638	ug/l	99
28) Bromochloromethane	4.885	49	289548	171.451	ug/l	99
29) Tetrahydrofuran	4.989	42	616374	774.811	ug/l	100
30) Chloroform	5.080	83	591709	151.367	ug/l	99
31) Cyclohexane	5.458	56	456119	141.570	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	506654	152.802	ug/l	99
36) 1,1-Dichloropropene	5.684	75	386601	143.556	ug/l	99
37) Ethyl Acetate	4.702	43	444461	154.141	ug/l	99
38) Carbon Tetrachloride	5.666	117	427263	146.021	ug/l	98
39) Methylcyclohexane	7.367	83	456694	149.395	ug/l	98
40) Benzene	6.025	78	1202846	147.057	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
 Data File : VX047590.D
 Acq On : 16 Sep 2025 12:01
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Sep 22 01:52:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.904	41	223218	149.951	ug/l	99
42) 1,2-Dichloroethane	6.074	62	455978	146.515	ug/l	99
43) Isopropyl Acetate	6.324	43	724888	153.283	ug/l	99
44) Trichloroethene	7.111	130	299673	151.542	ug/l	96
45) 1,2-Dichloropropane	7.415	63	316720	151.219	ug/l	99
46) Dibromomethane	7.568	93	228550	149.954	ug/l	97
47) Bromodichloromethane	7.806	83	483438	153.710	ug/l	98
48) Methyl methacrylate	7.678	41	367931	156.194	ug/l	99
49) 1,4-Dioxane	7.647	88	82739	3475.565	ug/l	98
51) 4-Methyl-2-Pentanone	8.562	43	2008063	765.849	ug/l	100
52) Toluene	8.702	92	739332	146.712	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	503912	157.076	ug/l	100
54) cis-1,3-Dichloropropene	8.354	75	530566	154.732	ug/l	99
55) 1,1,2-Trichloroethane	9.135	97	289103	148.781	ug/l	99
56) Ethyl methacrylate	9.104	69	490623	160.773	ug/l	97
57) 1,3-Dichloropropane	9.293	76	502121	150.426	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.232	63	1202206	749.661	ug/l	99
59) 2-Hexanone	9.415	43	1412417	750.035	ug/l	99
60) Dibromochloromethane	9.506	129	347748	155.318	ug/l	98
61) 1,2-Dibromoethane	9.592	107	302715	152.949	ug/l	99
64) Tetrachloroethene	9.263	164	227931	144.590	ug/l	98
65) Chlorobenzene	10.061	112	828239	150.763	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.147	131	291433	153.733	ug/l	98
67) Ethyl Benzene	10.177	91	1431554	151.026	ug/l	99
68) m/p-Xylenes	10.287	106	1059776	300.381	ug/l	99
69) o-Xylene	10.628	106	521840	152.920	ug/l	99
70) Styrene	10.640	104	917532	153.449	ug/l	99
71) Bromoform	10.781	173	229683	162.262	ug/l	99
73) Isopropylbenzene	10.945	105	1321409	151.477	ug/l	99
74) N-amyl acetate	10.829	43	636879	160.099	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.195	83	403906	153.035	ug/l	99
76) 1,2,3-Trichloropropane	11.226	75	470623	202.576	ug/l	89
77) Bromobenzene	11.183	156	328336	152.818	ug/l	99
78) n-propylbenzene	11.287	91	1550316	150.337	ug/l	99
79) 2-Chlorotoluene	11.348	91	939707	148.781	ug/l	99
80) 1,3,5-Trimethylbenzene	11.433	105	1091593	152.306	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.000	75	145921	160.176	ug/l	97
82) 4-Chlorotoluene	11.439	91	1127233	151.112	ug/l	100
83) tert-Butylbenzene	11.701	119	1106532	150.819	ug/l	99
84) 1,2,4-Trimethylbenzene	11.738	105	1106914	152.336	ug/l	100
85) sec-Butylbenzene	11.872	105	1319819	151.448	ug/l	99
86) p-Isopropyltoluene	11.994	119	1103210	149.530	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	605123	151.633	ug/l	100
88) 1,4-Dichlorobenzene	12.024	146	613617	149.814	ug/l	99
89) n-Butylbenzene	12.317	91	1032357	151.222	ug/l	100
90) Hexachloroethane	12.518	117	211472	163.258	ug/l	98
91) 1,2-Dichlorobenzene	12.317	146	582565	153.227	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.920	75	89415	167.332	ug/l	96
93) 1,2,4-Trichlorobenzene	13.567	180	387652	156.884	ug/l	98
94) Hexachlorobutadiene	13.707	225	121791	145.110	ug/l	99
95) Naphthalene	13.756	128	1229363	163.388	ug/l	100
96) 1,2,3-Trichlorobenzene	13.939	180	369382	160.704	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047590.D
Acq On : 16 Sep 2025 12:01
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Quant Time: Sep 22 01:52:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091625\
Data File : VX047590.D
Acq On : 16 Sep 2025 12:01
Operator : JC/MD
Sample : VSTDIC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Quant Time: Sep 22 01:52:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:168 Resp: 153177
Ion Ratio Lower Upper
168 100
99 61.1 48.8 73.2



#2
Dichlorodifluoromethane
Concen: 142.586 ug/l
RT: 1.173 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 85 Resp: 226166
Ion Ratio Lower Upper
85 100
87 31.8 15.8 47.3









#3
Chloromethane
Concen: 142.593 ug/l
RT: 1.301 min Scan# 35
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 297259
Ion Ratio Lower Upper
50 100
52 32.2 26.1 39.1



#4
Vinyl Chloride
Concen: 148.140 ug/l
RT: 1.380 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 62 Resp: 302303
Ion Ratio Lower Upper
62 100
64 32.1 24.9 37.3









#5
Bromomethane
Concen: 114.019 ug/l
RT: 1.605 min Scan# 81
Delta R.T. -0.012 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 94 Resp: 147541
Ion Ratio Lower Upper
94 100
96 93.4 76.6 115.0



#6
Chloroethane
Concen: 143.652 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 64 Resp: 195927
Ion Ratio Lower Upper
64 100
66 31.9 26.1 39.1









#7
 Trichlorofluoromethane
 Concen: 149.545 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Tgt Ion:101 Resp: 453116
 Ion Ratio Lower Upper
 101 100
 103 63.9 51.4 77.2



#8
 Diethyl Ether
 Concen: 151.684 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01

Tgt Ion: 74 Resp: 188006
 Ion Ratio Lower Upper
 74 100
 45 105.5 52.3 156.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 151.264 ug/l
RT: 2.331 min Scan# 205
Delta R.T. -0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 267995
Ion Ratio Lower Upper
101 100
85 43.9 35.5 53.3
151 73.2 59.4 89.0



#10
Methyl Iodide
Concen: 147.384 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:142 Resp: 407242
Ion Ratio Lower Upper
142 100
127 47.3 37.2 55.8
141 14.3 11.4 17.0





#11

Tert butyl alcohol

Concen: 795.166 ug/l

RT: 2.959 min Scan# 304

Delta R.T. 0.018 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 59 Resp: 215927

Ion Ratio Lower Upper

59 100

57 10.2 8.3 12.5



#12

1,1-Dichloroethene

Concen: 150.388 ug/l

RT: 2.325 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 96 Resp: 273692


Ion Ratio Lower Upper

96 100

61 166.1 131.0 196.6

98 64.0 47.7 71.5








#13
Acrolein
Concen: 894.203 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 223763
Ion Ratio Lower Upper
56 100
55 69.6 57.2 85.8



#14
Allyl chloride
Concen: 147.727 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 41 Resp: 554779
Ion Ratio Lower Upper
41 100
39 67.3 54.7 82.1
76 32.4 25.7 38.5









#15
Acrylonitrile
Concen: 786.201 ug/l
RT: 3.063 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 53 Resp: 790790
Ion Ratio Lower Upper
53 100
52 82.6 66.4 99.6
51 36.1 29.5 44.3



#16
Acetone
Concen: 647.201 ug/l
RT: 2.380 min Scan# 212
Delta R.T. 0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 43 Resp: 686119
Ion Ratio Lower Upper
43 100
58 30.0 23.4 35.0









#17
Carbon Disulfide
Concen: 143.412 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 76 Resp: 714503
Ion Ratio Lower Upper
76 100
78 9.0 6.8 10.2



#18
Methyl Acetate
Concen: 171.548 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 43 Resp: 404736
Ion Ratio Lower Upper
43 100
74 22.1 17.4 26.2









#19
Methyl tert-butyl Ether
Concen: 152.826 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 1015295
Ion Ratio Lower Upper
73 100
57 21.9 17.4 26.2



#20
Methylene Chloride
Concen: 144.398 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 84 Resp: 324032
Ion Ratio Lower Upper
84 100
49 129.4 108.6 162.8
51 39.9 32.6 48.8
86 63.8 52.0 78.0









#21
trans-1,2-Dichloroethene
Concen: 148.833 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 291616
Ion Ratio Lower Upper
96 100
61 144.0 116.2 174.2
98 64.3 49.3 73.9



#22
Diisopropyl ether
Concen: 150.551 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 45 Resp: 1095726
Ion Ratio Lower Upper
45 100
43 104.7 81.0 121.4
87 26.2 20.3 30.5
59 11.1 9.0 13.4









#23
 Vinyl Acetate
 Concen: 773.364 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 4504849
 Ion Ratio Lower Upper
 43 100
 86 10.0 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 151.776 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01

Tgt Ion: 63 Resp: 587129
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.8 11.2
 100 4.3 2.3 6.9









#25
2-Butanone
Concen: 736.502 ug/l
RT: 4.544 min Scan# 566
Delta R.T. 0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 973938
Ion Ratio Lower Upper
43 100
72 24.6 20.1 30.1



#26
2,2-Dichloropropane
Concen: 151.414 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 77 Resp: 486890
Ion Ratio Lower Upper
77 100
97 22.1 10.9 32.7





#27

cis-1,2-Dichloroethene

Concen: 152.638 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 366262

Ion Ratio Lower Upper

96 100

61 149.8 0.0 303.2

98 64.6 0.0 127.8



#28

Bromochloromethane

Concen: 171.451 ug/l

RT: 4.885 min Scan# 623

Delta R.T. -0.006 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 49 Resp: 289548


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.2

130 73.2 57.8 86.8





#29

Tetrahydrofuran

Concen: 774.811 ug/l

RT: 4.989 min Scan# 640

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 616374

Ion Ratio Lower Upper

42 100

72 42.8 34.2 51.2

71 39.9 31.9 47.9



#30

Chloroform

Concen: 151.367 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.000 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01


Tgt Ion: 83 Resp: 591709

Ion Ratio Lower Upper

83 100

85 65.2 51.8 77.6









#31
Cyclohexane
Concen: 141.570 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 456119
Ion Ratio Lower Upper
56 100
69 30.5 23.8 35.8
84 81.5 64.6 97.0



#32
1,1,1-Trichloroethane
Concen: 152.802 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 97 Resp: 506654
Ion Ratio Lower Upper
97 100
99 63.6 50.6 76.0
61 44.2 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 162.673 ug/l

RT: 5.940 min Scan# 796

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 65 Resp: 396627

Ion Ratio Lower Upper

65 100

67 52.3 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.745 min Scan# 928

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 114 Resp: 274804


Ion Ratio Lower Upper

114 100

63 22.3 0.0 44.0

88 16.8 0.0 34.2





#35

Dibromofluoromethane

Concen: 164.802 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 303728

Ion Ratio Lower Upper

113 100

111 102.4 80.9 121.3

192 18.1 14.2 21.4



#36

1,1-Dichloropropene

Concen: 143.556 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 75 Resp: 386601


Ion Ratio Lower Upper

75 100

110 33.9 16.7 50.1

77 30.6 24.3 36.5









#37
Ethyl Acetate
Concen: 154.141 ug/l
RT: 4.702 min Scan# 593
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 444461
Ion Ratio Lower Upper
43 100
61 12.1 9.8 14.8
70 9.5 7.9 11.9



#38
Carbon Tetrachloride
Concen: 146.021 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 117 Resp: 427263
Ion Ratio Lower Upper
117 100
119 93.3 73.4 110.2
121 29.0 23.8 35.6









#39
Methylcyclohexane
Concen: 149.395 ug/l
RT: 7.367 min Scan# 1030
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 83 Resp: 456694
Ion Ratio Lower Upper
83 100
55 80.7 66.4 99.6
98 48.3 38.1 57.1



#40
Benzene
Concen: 147.057 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 78 Resp: 1202846
Ion Ratio Lower Upper
78 100
77 23.9 18.8 28.2








#41
Methacrylonitrile
Concen: 149.951 ug/l
RT: 4.904 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 41 Resp: 223218
Ion Ratio Lower Upper
41 100
39 54.5 43.3 64.9
67 74.6 58.6 87.8
52 32.0 26.0 39.0



#42
1,2-Dichloroethane
Concen: 146.515 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 62 Resp: 455978
Ion Ratio Lower Upper
62 100
98 8.7 0.0 17.0









#43
Isopropyl Acetate
Concen: 153.283 ug/l
RT: 6.324 min Scan# 859
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 724888
Ion Ratio Lower Upper
43 100
61 19.4 15.2 22.8
87 12.3 9.6 14.4



#44
Trichloroethene
Concen: 151.542 ug/l
RT: 7.111 min Scan# 988
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:130 Resp: 299673
Ion Ratio Lower Upper
130 100
95 101.1 0.0 210.8





#45

1,2-Dichloropropane

Concen: 151.219 ug/l

RT: 7.415 min Scan# 1038

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 63 Resp: 316720

Ion Ratio Lower Upper

63 100

65 31.5 24.7 37.1



#46

Dibromomethane

Concen: 149.954 ug/l

RT: 7.568 min Scan# 1063

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 93 Resp: 228550


Ion Ratio Lower Upper

93 100

95 84.2 65.0 97.6

174 92.5 71.6 107.4









#47
Bromodichloromethane
Concen: 153.710 ug/l
RT: 7.806 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 483438
Ion Ratio Lower Upper
83 100
85 63.9 52.5 78.7
127 7.9 6.9 10.3



#48
Methyl methacrylate
Concen: 156.194 ug/l
RT: 7.678 min Scan# 1081
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 41 Resp: 367931
Ion Ratio Lower Upper
41 100
69 77.2 61.4 92.2
39 56.5 45.8 68.6





#49

1,4-Dioxane

Concen: 3475.565 ug/l

RT: 7.647 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150




Tgt Ion: 88 Resp: 82739

Ion Ratio Lower Upper

88 100

43 38.5 30.2 45.2

58 73.0 59.5 89.3



#50

Toluene-d8

Concen: 163.253 ug/l

RT: 8.635 min Scan# 1238

Delta R.T. -0.000 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01

Tgt Ion: 98 Resp: 1041079

Ion Ratio Lower Upper

98 100

100 67.3 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 765.849 ug/l
RT: 8.562 min Scan# 1226
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 2008063
Ion Ratio Lower Upper
43 100
58 38.0 30.3 45.5



#52
Toluene
Concen: 146.712 ug/l
RT: 8.702 min Scan# 1249
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 92 Resp: 739332
Ion Ratio Lower Upper
92 100
91 169.5 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 157.076 ug/l

RT: 8.964 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 75 Resp: 503912

Ion Ratio Lower Upper

75 100

77 31.6 25.4 38.2



#54

cis-1,3-Dichloropropene

Concen: 154.732 ug/l

RT: 8.354 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion: 75 Resp: 530566


Ion Ratio Lower Upper

75 100

77 32.0 25.8 38.6

39 52.3 41.4 62.2









#55
1,1,2-Trichloroethane
Concen: 148.781 ug/l
RT: 9.135 min Scan# 1320
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:	97	Resp:	289103
Ion	Ratio	Lower	Upper
97	100		
83	86.1	69.6	104.4
85	56.8	45.0	67.6
99	62.6	49.6	74.4



#56
Ethyl methacrylate
Concen: 160.773 ug/l
RT: 9.104 min Scan# 1315
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:	69	Resp:	490623
Ion	Ratio	Lower	Upper
69	100		
41	70.5	58.6	87.8
39	38.2	31.5	47.3





#57

1,3-Dichloropropane

Concen: 150.426 ug/l

RT: 9.293 min Scan# 1171

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 502121

Ion Ratio Lower Upper

76 100

78 31.9 26.2 39.4



#58

2-Chloroethyl Vinyl ether

Concen: 749.661 ug/l

RT: 8.232 min Scan# 1172

Delta R.T. 0.006 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01


Tgt Ion: 63 Resp: 1202206

Ion Ratio Lower Upper

63 100

106 27.7 21.9 32.9









#59
2-Hexanone
Concen: 750.035 ug/l
RT: 9.415 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1412417
Ion Ratio Lower Upper
43 100
58 52.9 26.1 78.2



#60
Dibromochloromethane
Concen: 155.318 ug/l
RT: 9.506 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:129 Resp: 347748
Ion Ratio Lower Upper
129 100
127 78.7 38.4 115.2








#61
1,2-Dibromoethane
Concen: 152.949 ug/l
RT: 9.592 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 302715
Ion Ratio Lower Upper
107 100
109 93.5 75.3 112.9



#62
4-Bromofluorobenzene
Concen: 165.020 ug/l
RT: 11.067 min Scan# 1637
Delta R.T. 0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 95 Resp: 399386
Ion Ratio Lower Upper
95 100
174 70.7 0.0 141.6
176 68.5 0.0 138.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 241818
Ion Ratio Lower Upper
117 100
82 58.1 47.8 71.6
119 32.2 25.2 37.8



#64
Tetrachloroethene
Concen: 144.590 ug/l
RT: 9.263 min Scan# 1341
Delta R.T. 0.006 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:164 Resp: 227931
Ion Ratio Lower Upper
164 100
166 127.8 100.2 150.2
129 99.0 81.8 122.6
131 96.4 78.6 118.0





#65

Chlorobenzene

Concen: 150.763 ug/l

RT: 10.061 min Scan# 1472

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:112 Resp: 828239

Ion Ratio Lower Upper

112 100

114 31.8 25.1 37.7



#66

1,1,1,2-Tetrachloroethane

Concen: 153.733 ug/l

RT: 10.147 min Scan# 1486

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion:131 Resp: 291433


Ion Ratio Lower Upper

131 100

133 96.7 47.3 142.0

119 64.8 32.0 96.0









#67
Ethyl Benzene
Concen: 151.026 ug/l
RT: 10.177 min Scan# 1491
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 1431554
Ion Ratio Lower Upper
91 100
106 30.0 24.3 36.5



#68
m/p-Xylenes
Concen: 300.381 ug/l
RT: 10.287 min Scan# 1509
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:106 Resp: 1059776
Ion Ratio Lower Upper
106 100
91 208.0 167.8 251.8





#69

o-Xylene

Concen: 152.920 ug/l

RT: 10.628 min Scan# 1564

Delta R.T. 0.006 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:106 Resp: 521840

Ion Ratio Lower Upper

106 100

91 220.7 111.1 333.4



#70

Styrene

Concen: 153.449 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion:104 Resp: 917532

Ion Ratio Lower Upper

104 100

78 52.4 43.1 64.7

103 54.0 43.1 64.7





#71

Bromoform

Concen: 162.262 ug/l

RT: 10.781 min Scan# 1590

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150




Tgt Ion:173 Resp: 229683

Ion Ratio Lower Upper

173 100

175 48.7 24.9 74.6

254 0.1 0.1 0.1



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion:152 Resp: 114743

Ion Ratio Lower Upper

152 100

115 136.9 44.9 134.5#

150 212.9 0.0 352.0








#73
Isopropylbenzene
Concen: 151.477 ug/l
RT: 10.945 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:105 Resp: 1321409
Ion Ratio Lower Upper
105 100
120 26.2 12.8 38.4



#74
N-ethyl acetate
Concen: 160.099 ug/l
RT: 10.829 min Scan# 1598
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 43 Resp: 636879
Ion Ratio Lower Upper
43 100
70 40.4 32.0 48.0
55 22.9 18.9 28.3
61 23.6 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 153.035 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 403906

Ion Ratio Lower Upper

83 100

131 10.0 5.0 14.9

85 64.6 31.6 95.0



#76

1,2,3-Trichloropropane

Concen: 202.576 ug/l

RT: 11.226 min Scan# 1663

Delta R.T. 0.006 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01


Tgt Ion: 75 Resp: 470623

Ion Ratio Lower Upper

75 100

77 30.8 18.8 56.4





#77

Bromobenzene

Concen: 152.818 ug/l

RT: 11.183 min Scan# 1656

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:156 Resp: 328336

Ion Ratio Lower Upper

156 100

77 178.2 89.7 269.1

158 97.5 47.9 143.6



#78

n-propylbenzene

Concen: 150.337 ug/l

RT: 11.287 min Scan# 1673

Delta R.T. -0.000 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01


Tgt Ion: 91 Resp: 1550316

Ion Ratio Lower Upper

91 100

120 22.7 11.1 33.1









#79
2-Chlorotoluene
Concen: 148.781 ug/l
RT: 11.348 min Scan# 1683
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 939707
Ion Ratio Lower Upper
91 100
126 32.7 16.1 48.3



#80
1,3,5-Trimethylbenzene
Concen: 152.306 ug/l
RT: 11.433 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 105 Resp: 1091593
Ion Ratio Lower Upper
105 100
120 47.4 23.6 70.8









#81
trans-1,4-Dichloro-2-butene
Concen: 160.176 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 145921
Ion Ratio Lower Upper
75 100
53 100.8 83.4 125.2
89 46.3 36.3 54.5



#82
4-Chlorotoluene
Concen: 151.112 ug/l
RT: 11.439 min Scan# 1698
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 91 Resp: 1127233
Ion Ratio Lower Upper
91 100
126 28.5 14.2 42.6





#83

tert-Butylbenzene

Concen: 150.819 ug/l

RT: 11.701 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:119 Resp: 1106532

Ion Ratio Lower Upper

119 100

91 58.7 29.6 88.8

134 22.3 11.0 33.0



#84

1,2,4-Trimethylbenzene

Concen: 152.336 ug/l

RT: 11.738 min Scan# 1747

Delta R.T. 0.006 min

Lab File: VX047590.D




Acq: 16 Sep 2025 12:01


Tgt Ion:105 Resp: 1106914

Ion Ratio Lower Upper

105 100

120 44.4 22.1 66.1









#85
 sec-Butylbenzene
 Concen: 151.448 ug/l
 RT: 11.872 min Scan# 1769
 Delta R.T. -0.000 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 1319819
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.5



#86
 p-Isopropyltoluene
 Concen: 149.530 ug/l
 RT: 11.994 min Scan# 1789
 Delta R.T. -0.000 min
 Lab File: VX047590.D
 Acq: 16 Sep 2025 12:01

Tgt Ion:119 Resp: 1103210
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.6 37.8
 91 25.3 12.6 37.6









#87
1,3-Dichlorobenzene
Concen: 151.633 ug/l
RT: 11.951 min Scan# 1782
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 605123
Ion Ratio Lower Upper
146 100
111 42.8 21.5 64.5
148 63.8 32.0 96.0



#88
1,4-Dichlorobenzene
Concen: 149.814 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion:146 Resp: 613617
Ion Ratio Lower Upper
146 100
111 41.0 20.8 62.4
148 63.2 31.7 95.1





#89

n-Butylbenzene

Concen: 151.222 ug/l

RT: 12.317 min Scan# 1842

Delta R.T. 0.006 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :





MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 1032357

Ion	Ratio	Lower	Upper
91	100		
92	55.2	27.6	82.7
134	25.2	12.6	37.8



#90

Hexachloroethane

Concen: 163.258 ug/l

RT: 12.518 min Scan# 1875




Delta R.T. -0.000 min


Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Tgt Ion: 117 Resp: 211472

Ion	Ratio	Lower	Upper
117	100		
201	66.6	34.2	102.5









#91
1,2-Dichlorobenzene
Concen: 153.227 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 582565
Ion Ratio Lower Upper
146 100
111 44.5 22.2 66.6
148 63.5 31.8 95.4



#92
1,2-Dibromo-3-Chloropropane
Concen: 167.332 ug/l
RT: 12.920 min Scan# 1941
Delta R.T. -0.000 min
Lab File: VX047590.D
Acq: 16 Sep 2025 12:01

Tgt Ion: 75 Resp: 89415
Ion Ratio Lower Upper
75 100
155 81.3 37.7 113.1
157 100.6 49.8 149.4





#93

1,2,4-Trichlorobenzene

Concen: 156.884 ug/l

RT: 13.567 min Scan# 2047

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 387652

Ion Ratio Lower Upper

180 100

182 95.3 46.6 139.8

145 33.5 16.4 49.1



#94

Hexachlorobutadiene

Concen: 145.110 ug/l

RT: 13.707 min Scan# 2070

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01




Tgt Ion:225 Resp: 121791


Ion Ratio Lower Upper

225 100

223 64.2 31.7 95.1

227 62.8 32.0 96.0





#95

Naphthalene

Concen: 163.388 ug/l

RT: 13.756 min Scan# 2078

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:128 Resp: 1229363

Ion Ratio Lower Upper

128 100

127 12.7 10.2 15.2

129 10.9 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 160.704 ug/l

RT: 13.939 min Scan# 2108

Delta R.T. -0.000 min

Lab File: VX047590.D

Acq: 16 Sep 2025 12:01

Tgt Ion:180 Resp: 369382

Ion Ratio Lower Upper

180 100

182 94.6 47.6 142.8

145 34.9 17.9 53.7

